



## ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

Statistical analysis and clusterization of environmental parameters measurements from co-located sensors

*Original*

Statistical analysis and clusterization of environmental parameters measurements from co-located sensors / Coppa, Graziano; Pennecchi, FRANCESCA ROMANA; Soncini, Manuele. - (2022). (Intervento presentato al convegno TECO-2022 tenutosi a Parigi, Francia nel 10-13 ottobre 2022).

*Availability:*

This version is available at: 11696/75099 since: 2022-12-21T12:38:15Z

*Publisher:*

*Published*

DOI:

*Terms of use:*

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

*Publisher copyright*

(Article begins on next page)



# Statistical analysis and clusterization of environmental parameters measurements from co-located sensors

## Abstract

One of the most difficult things in metrology for the environment is the **evaluation of measurement uncertainties**. **Air temperature sensors are influenced by many external factors**, such as solar radiation – both direct and diffuse – humidity, wind speed and direction. Laboratory calibration is necessary but not sufficient for a comprehensive understanding of the behaviour of air temperature sensors while subject to the wide variety of environmental conditions they meet in normal operation.

This work follows up the one by **Coppa et al., (2021)**, where the evaluation of uncertainties of commercially available air temperature sensors and shields was computed by a dedicated experiment, where 7 sensors were co-located for 2 months. In that work, **uncertainties were evaluated simply by computing standard deviations of temperature differences between each sensor and one taken as a reference**.


## The road siting effect evaluation experiment



Sensor array during normal phase of the experiment.



Co-located sensors on the experimental field site.




Distributions of temperature differences between each sensor and a reference during co-location of the measurement stations. Blue vertical dashed lines represent means.

Coppa, Quarello, Steeneveld, Jandric and Merlone, 2021, *IntJClim*, 41: 3705– 3724 DOI: 10.1002/joc.7044


The present work provides a more refined uncertainty evaluation by a statistical method, through **machine learning and clusterization**, in order to **evaluate uncertainties in dependence on the environmental conditions**.

Temperature differences during the co-location experiment have been run through an **unsupervised k-means algorithm** that revealed several clusters: in particular, **temperature differences are lower, thus leading to smaller uncertainty**, under two very different environmental conditions: the first, **at nights with very low winds**; the second **during high-radiation and high-wind days**.


## Unsupervised k-means clustering algorithm



Numerosity of the meteorological observations in the 4 clusters identified by the k-means algorithm.



Distributions of meteorological variables in the 4 clusters identified by the k-means algorithm.



Representation of 4 clusters in a simplified 3D space of the meteorological parameters.

**Temperature differences can be corrected** by subtracting the **cluster centroid** value from each observation, while **uncertainties can be taken** as the **standard deviations of each cluster distributions**, then associated to each sensor+shield system depending on the specific environmental conditions (evaluated by ancillary measurements). This methodology **reduces the overall uncertainty** and **improves the level of detail** about sensors behaviour knowledge.

	$\Delta T$ (°C)	Rad (W/m <sup>2</sup> )	RH (%)	Wind speed (m/s)
Cluster 0	0.047	7	94.0	0.71
Cluster 1	0.352	192	71.0	0.73
Cluster 2	0.109	207	54.1	1.24
Cluster 3	0.051	106	49.9	3.11

Cluster centroids of the 4 different parameters.

	$\sigma_{\Delta T}$ (°C)	Variation
Cluster 0	0.045	43%
Cluster 1	0.120	115%
Cluster 2	0.070	67%
Cluster 3	0.055	53%

Standard deviations of  $\Delta T$  measurements and variation over previous evaluation.

## Conclusions

- A way to refine the evaluation of measurement uncertainties in air temperature, is presented.
- A clustering algorithm to group together measurements taken in similar conditions is used
- Centroids as bias corrections
- Standard deviations of cluster as uncertainties
- Uncertainties for the largest part of measurements are reduced

## Future work

- Extend to longer time periods (whole year)
- Explore other clustering algorithms