



ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

Statistical analysis and clusterization of environmental parameters measurements from co-located sensors

Original

Statistical analysis and clusterization of environmental parameters measurements from co-located sensors / Coppa, Graziano; Pennechi, FRANCESCA ROMANA; Soncini, Manuele. - (2022). (Intervento presentato al convegno TECO-2022 tenutosi a Parigi, Francia nel 10-13 ottobre 2022).

Availability:

This version is available at: 11696/75099 since: 2022-12-21T12:38:15Z

Publisher:

Published

DOI:

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

Statistical analysis and clusterization of environmental parameters measurements from co-located sensors

Abstract

One of the most difficult things in metrology for the environment is the **evaluation of measurement uncertainties**. Air temperature sensors are **influenced by many external factors**, such as solar radiation – both direct and diffuse – humidity, wind speed and direction. Laboratory calibration is necessary but not sufficient for a comprehensive understanding of the behaviour of air temperature sensors while subject to the wide variety of environmental conditions they meet in normal operation.

This work follows up the one by **Coppa et al., (2021)**, where the evaluation of uncertainties of commercially available air temperature sensors and shields was computed by a dedicated experiment, where 7 sensors were co-located for 2 months. In that work, **uncertainties were evaluated simply by computing standard deviations of temperature differences between each sensor and one taken as a reference**.


The road siting effect evaluation experiment



Sensor array during normal phase of the experiment.



Co-located sensors on the experimental field site.




Distributions of temperature differences between each sensor and a reference during co-location of the measurement stations. Blue vertical dashed lines represent means.

Coppa, Quarello, Steeneveld, Jandric and Merlone, 2021, *IntJCLim*, 41: 3705– 3724 DOI: 10.1002/joc.7044


The present work provides a more refined uncertainty evaluation by a statistical method, through **machine learning and clusterization**, in order to **evaluate uncertainties in dependence on the environmental conditions**.

Temperature differences during the co-location experiment have been run through an **unsupervised k-means algorithm** that revealed several clusters: in particular, **temperature differences are lower, thus leading to smaller uncertainty**, under two very different environmental conditions: the first, **at nights with very low winds**; the second **during high-radiation and high-wind days**.


Unsupervised k-means clustering algorithm



Numerosity of the meteorological observations in the 4 clusters identified by the k-means algorithm.



Distributions of meteorological variables in the 4 clusters identified by the k-means algorithm.



Representation of 4 clusters in a simplified 3D space of the meteorological parameters.

Temperature differences can be corrected by subtracting the **cluster centroid** value from each observation, while **uncertainties can be taken** as the **standard deviations of each cluster distributions**, then associated to each sensor+shield system depending on the specific environmental conditions (evaluated by ancillary measurements). This methodology **reduces the overall uncertainty** and **improves the level of detail** about sensors behaviour knowledge.

	ΔT (°C)	Rad (W/m ²)	RH (%)	Wind speed (m/s)
Cluster 0	0.047	7	94.0	0.71
Cluster 1	0.352	192	71.0	0.73
Cluster 2	0.109	207	54.1	1.24
Cluster 3	0.051	106	49.9	3.11

Cluster centroids of the 4 different parameters.

	$\sigma_{\Delta T}$ (°C)	Variation
Cluster 0	0.045	43%
Cluster 1	0.120	115%
Cluster 2	0.070	67%
Cluster 3	0.055	53%

Standard deviations of ΔT measurements and variation over previous evaluation.

Conclusions

- A way to refine the evaluation of measurement uncertainties in air temperature, is presented.
- A clustering algorithm to group together measurements taken in similar conditions is used
- Centroids as bias corrections
- Standard deviations of cluster as uncertainties
- Uncertainties for the largest part of measurements are reduced

Future work

- Extend to longer time periods (whole year)
- Explore other clustering algorithms