

ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

The kilogram: inertial or gravitational mass?

Original The kilogram: inertial or gravitational mass? / Mana, G; Schlamminger, S In: METROLOGIA ISSN 0026-1394 59:4(2022), p. 043001. [10.1088/1681-7575/ac7ca7]
Availability: This version is available at: 11696/75440 since: 2023-02-08T13:42:25Z
Publisher: IOP Publishing Ltd
Published DOI:10.1088/1681-7575/ac7ca7
Terms of use:
This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

(Article begins on next page)

Publisher copyright



LETTER TO THE EDITOR • OPEN ACCESS

The kilogram: inertial or gravitational mass?

To cite this article: G Mana and S Schlamminger 2022 Metrologia 59 043001

View the article online for updates and enhancements.

You may also like

- Self-attraction mapping and an update on local gravitational acceleration measurement in BIPM Kibble balance Shisong Li, Franck Bielsa, Adrien Kiss et al.
- The watt or Kibble balance: a technique for implementing the new SI definition of the unit of mass Ian A Robinson and Stephan Schlamminger
- Realization of the kilogram using the KRISS Kibble balance
 Dongmin Kim, MyeongHyeon Kim, Minky Seo et al.

Metrologia 59 (2022) 043001 (4pp)

The kilogram: inertial or gravitational mass?

G Mana^{1,2} and S Schlamminger^{3,*}

- ¹ INRIM—Istituto Nazionale di Ricerca Metrologica, Torino, Italy
- ² Dipartimento di Fisica, UNITO—Università di Torino, Torino, Italy
- ³ NIST—National Institute of Standard and Technology, 100 Bureau Drive, Gaithersburg, MD 20899, United States of America

E-mail: stephan.schlamminger@nist.gov

Received 16 January 2022, revised 24 June 2022 Accepted for publication 28 June 2022 Published 5 August 2022



Abstract

With the redefinition of the international system of units, the value of the Planck constant was fixed, similarly to the values of the unperturbed ground state hyperfine transition frequency of the ¹³³Cs atom and speed of light in vacuum. Theoretically and differently from the past, the kilogram is now explicitly defined as the unit of inertial mass. Experimentally, the kilogram is realized by atom count or the Kibble balance. We show that only the former method measures the inertial mass without assuming the universality of free fall. Under ordinary circumstances, the results obtained by the Kibble balance require the equivalence of inertial and gravitational mass. Therefore, the agreement between the two measures can be interpreted as a test of the equivalence principle.

Keywords: kilogram, equivalence principle, inertial mass, gravitational mass

1. Introduction

A constant that is never made explicit links inertial to gravitational mass of all matter and energy. The inertial mass, m_i , determines the force required to accelerate an object by a given rate, $m_i = F/a$. The gravitational mass, m_g , determines the force in Newton's law of universal gravitation, $F = Gm_gM_g/r^2$ and plays a role similar to the charge in Coulomb's law.

Since Newton's unification of the Earth and celestial mechanics, the equivalence principle states that they are the same quantity. It implies the universality of free fall: in a gravitational field, locally, all bodies fall with the same acceleration, independently of their composition. Accepting this principle, the constant linking the gravitational and inertial masses must be dimensionless and can be conveniently set equal to one.

The equivalence principle is an axiom of physics. Historically, the embrace of its validity led to the development of Newton's theory of gravity [1], as well as of its successor, gen-

* Author to whom any correspondence should be addressed.

Original content from this work may be used under the terms of the Creative Commons Attribution 4.0 licence. Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.

eral relativity [2]. Upper limits to the principle violation must be established experimentally. Suppose that, for a material X, the ratio of gravitational to inertial masses is $m_{\rm g}(X)/m_{\rm i}(X)=1+\eta(X)$. Then, by comparing the free-fall acceleration of two different materials X and Y, the statement

$$\frac{a(X)}{a(Y)} = \frac{1 + \eta(X)}{1 + \eta(Y)} \approx 1 + \eta(X) - \eta(Y) \tag{1}$$

can be made on the difference of the η values.

The experimental tests date back to Galilei [3] and potentially to earlier. A large sensitivity advance was made when von Eötvös [4] realized that the sought difference, instead of being calculated from two large measurement results, a(X) and a(Y), as Galilei did, can be experimentally obtained via a null experiment, for example, by using a torsion balance. Such an experiment produces null if $\Delta \eta(X, Y) := \eta(X) - \eta(Y)$ is smaller than its sensitivity. Since then, large reductions of the $\Delta \eta$ upper limit have been reported [5–8].

It is worth noting that all tests involve at least two materials, here symbolised by X and Y, and deliver only a statement about an upper limit of the specific $\Delta \eta(X,Y)$

In 1901, the 3rd Conférence Générale des Poids et Mesures declared that the unit of mass is equal to the mass of the

Metrologia 59 (2022) 043001 G Mana and S Schlamminger

international prototype of the kilogram, a Pt–Ir artefact. This definition, consistent with the equivalence principle, did not distinguish between the inertial and gravitational masses. However, disseminating the kilogram by balances, we compared gravitational masses. In this sense, the quantity that was traced back to the Pt–Ir prototype was the gravitational mass.

The 2019 redefinition of the international system of units changed this state of affairs [9]. The unit of mass is now traced back to the stipulated values of the unperturbed ground state hyperfine transition frequency of the ¹³³Cs atom, $\nu_{\rm Cs}$, speed of light in vacuum, c, and Planck constant, h. Therefore, the mass defect between the two hyperfine ground-state levels of the ¹³³Cs atom is exactly $\Delta m_{\rm Cs} = h\nu_{\rm Cs}/c^2$. The roots of $\Delta m_{\rm Cs}$ are the Einstein and Planck equations $E = mc^2$ and $E = h\nu$. In special relativity, the equality of energy and the inertial (rest) mass follows from the conservation law for the energy–momentum tensor [10], where gravity plays no role. Hence, $\Delta m_{\rm Cs}$ relates the inertial masses of the ¹³³Cs atom before and after the transition, and the kilogram is now the unit of the inertial mass.

The object of this short communication is to examine how the equivalence principle underlies the practical realizations of the kilogram via counting atoms and the Kibble balance.

2. Atom count

We conceptually describe how the atom counting method can be used to realize an inertial mass standard. The first realisation step is recoiling 133 Cs or 87 Rb atoms by photons in an atom interferometer to measure the ratios between their inertial masses and the Planck constant [11, 12]. Alternatively, one can derive the $m_{\rm e}/h$ ratio from the measured value of the Rydberg constant via hydrogen spectroscopy. The ratio comes into the Rydberg constant from the kinetic term of the hydrogenatom Hamiltonian; therefore, $m_{\rm e}$ stands for the inertial mass of the electron. These mass ratios fix the absolute scale of atomic (inertial) masses via relative mass spectrometry by Penning traps.

In the second step, the kilogram is realised by atom counting. To determine the count in practice, a 28 Si monocrystal is shaped as a quasi-perfect ball; the number $N_{\rm Si}$ of atoms in it is obtained from the measurement of the ball volume V and lattice parameter a_0 according to $8V/a_0^3$ where $a_0^3/8$ is the atom volume, and 8 is the number of atoms in the cubic unit cell.

Making reference, for instance, to the m_e/h quotient, the measurement equation is

$$\frac{m_{\rm i}(^{28}{\rm Si\ ball})}{h} = \frac{8V}{a_0^3} \frac{M(^{28}{\rm Si})}{M(e)} \frac{m_{\rm e}}{h},\tag{2}$$

where m_i (²⁸Si ball) is the ball's inertial mass and M(X) indicates the X's molar mass [13]. Since Si crystals are never perfect, mono-isotopic, and pure, (2) is corrected for the isotope abundances, impurities, and point defects (vacancies and interstitials). Also, the ball surface is characterised to correct for the oxide layer, adsorbed or absorbed water, and contaminants. In principle, one should take the mass defect associated with the

binding energy of the atoms into account, but this correction is negligible at the present level of accuracy.

3. Kibble balance

Tracing mass measurements back to the Planck constant by a Kibble balance does not exactly imply the realisation of an inertial mass. Conceptually, a Kibble balance compares the power $m_g(K)gv$ generated by a gravitational mass $m_g(K)$ falling with constant velocity v in a locally uniform gravitational field g, with the power $\mathcal{E}I$ dissipated by magnet-coil brake that would keep the mass motion uniform (\mathcal{E} and I are the electromotive force and eddy current).

In practice, the balance's measurement-equation,

$$m_{\rm g}({\rm K})gv = \mathcal{E}I,$$
 (3)

is assembled in two steps. Firstly, one measures the current I necessary to hold up the mass in the brake's magnetic field. Next, the electromotive force \mathcal{E} is measured at the ends of the brake coil when the mass moves with constant velocity v. The tie to the Planck constant is provided by two electrical quantum standards. $\mathcal{E} = nf/K_J$ is measured in terms of the Josephson constant $K_J = 2e/h$, where n is an integer, e is the elementary charge, and f is a frequency [14]. The current I is converted into a voltage (that is again measured against K_J) passing it through a resistor that is a know fraction of the von-Klitzing constant $R_K = h/e^2$.

The local gravitational field is determined by tracking a free-falling body, a corner-cube mirror, with a laser interferometer. Let us assume that $m_i(ff)$ and $m_g(ff)$ are the inertial and gravitational masses of the free falling body. It is

$$m_{\rm g}({\rm ff})g = m_{\rm i}({\rm ff})a,$$
 (4)

where a is the kinematic acceleration observed by measuring the traveled distance, $z(t) = z_0 + v_0 t + at^2/2$. A curve fitting procedure yields a, and g is obtained as

$$g = \frac{m_{\rm i}(\rm ff)}{m_{\rm o}(\rm ff)}a. \tag{5}$$

Alternatives to the widely used classical gravimeters are measurements using atom or neutron interferometry and Bloch's oscillations of cold atoms in an optical lattice, which employ freely falling neutrons or atoms [15–17].

A careful analysis of neutron interferometers by Littrell and co-workers [18] shows that the difference of the quantum-mechanical phase accumulated by neutrons travelling the interferometer, $\Delta\Phi$, scales like $gm_g(n)$. However, this phase difference is measured as a fraction of the de Broglie wavelength, $\lambda = h/p$, with $p = m_i(n)v$ and v the neutrons' velocity. Thus,

$$\Delta\Phi \propto \frac{m_{\rm g}(n)}{m_{\rm i}(n)}g\tag{6}$$

Similar reasoning can be applied to atom interferometers and Bloch's oscillations in an optical lattice. The only difference is that the observable is not the de Broglie wavelength but the velocity and momentum changes of the free-falling atoms Metrologia 59 (2022) 043001 G Mana and S Schlamminger

as probed by photon absorption [17, 19, 20]. This process is kinematical and, hence, only sensitive to the inertial mass m_i .

Eventually, regardless of whether the probe mass is a macroscopic body, a neutron, or an atom, the measured value of g is essentially given by (5). Therefore, by using (5) in (3), the mass value that is obtained by using a Kibble balance is

$$m_{\text{KB}}(K) := m_{\text{g}}(K) \frac{m_{\text{i}}(\text{ff})}{m_{\text{g}}(\text{ff})} = \frac{\mathcal{E}I}{av},$$
 (7)

or, using the η symbol introduced earlier,

$$m_{\mathrm{KB}}(\mathrm{K}) := \frac{m_{\mathrm{g}}(\mathrm{K})}{1 + \eta(\mathrm{ff})} = \frac{1 + \eta(\mathrm{K})}{1 + \eta(\mathrm{ff})} m_{\mathrm{i}}(\mathrm{K}). \tag{8}$$

If the equivalence principle is assumed to hold true for the dropping object but not for the weighed one, i.e., $\eta(ff) = 0$ and $\eta(K) \neq 0$, then the Kibble balance gives the gravitational mass $m_g(K)$. Contrary, if $\eta(ff) = \eta(K)$, which could be achieved by dropping K, i.e., ff = K, then it gives the inertial mass $m_i(K)$. Lastly, for the case $\eta(ff) = \eta(K) = 0$, there is no distinction between the two types of masses.

A suggestion to eliminate the weighing dependence on gravity by operating the Kibble balance dynamically and measuring the inertial acceleration a of the test mass was made by Cabiati [21] and further investigated for a horizontal geometry by Kibble and Robinson [22]. In either of these arrangements, the inertial force $m_i(K)a$ substitutes for the gravitational force $m_g(K)g$ in (3), and inertial mass is measured without reference to the equivalence principle. As of today, Cabiati-type balances do not play a role in mass metrology and are not further discussed.

4. Conclusions

Contrary to the past, when balances disseminated gravitational masses, the kilogram is now per definition via h the unit of inertial mass. However, only the atom count determines the inertial mass of the kilogram realisation without any reference to the equivalence principle.

Conceptually, as discussed in the previous section, the quantity measured by the Kibble balance depends on the assumptions made on the equivalence principle. Recent experiments [23, 24] have shown the equivalence principle to hold at an uncertainty level that is many orders of magnitude smaller than the relative uncertainty obtained by Kibble balances. Hence, for all practical purposes, the question about what mass is measured is irrelevant.

Let us suppose that the mass of the same 28 Si ball is measured by both counting the atoms and the Kibble balance and that the mass values are found in agreement. The first key comparison of the kilogram realizations based on the fixed numerical value of the Planck constant [25, 26] is an embodiment of the supposed experiment, albeit in more than one step. The outcome is $m_{\rm KB}(^{28}{\rm Si~ball}) = m_{\rm i}(^{28}{\rm Si~ball})$. Combining this identity with (8), where $^{28}{\rm Si~ball}$ must substitute for K, yields

$$\eta(ff) - \eta(^{28}Si \text{ ball}) = 0 \tag{9}$$

and an upper limit of $\Delta\eta(\mathrm{ff},^{28}\mathrm{Si}\ \mathrm{ball})$ is obtained. Just as it was for Galileo's experiment, two large measurement results have been compared. It is, hence, not surprising that the sensitivity, approaching 10 $\mu\mathrm{g}\ \mathrm{kg}^{-1}$ at the best [26], is not competitive against that of null tests [23, 24], whose relative sensitivities reach 10^{-13} .

Comparing different kilogram realizations will never result in a competitive test of the equivalence principle. However, the musings in this article should not be dismissed for two special features of that comparison. Firstly, in contrast to null tests, the result reported here critically depends on the absolute weighing of a 1 kg body. Secondly, unlike Eötvös-like experiments, which compare gravitational and inertial accelerations, electromagnetic and gravitational accelerations are compared here.

Acknowledgments

Both authors want to thank the anonymous reviewers for their comments that helped clarify the text. GM received support from the Ministero dell'Istruzione, dell'Università e della Ricerca.

ORCID iDs

G Mana https://orcid.org/0000-0002-4109-7254 S Schlamminger https://orcid.org/0000-0002-9270-4018

References

- [1] Newton S I 1687 Newton's Principia: The Mathematical Principles of Natural Philosophy (New York: Daniel Adee) (Engl. transl. from 1846)
- [2] Einstein A 1915 Sitzungsberichte der Königlich Preußischen Akademie der Wissenschaften (Berlin: Verlag der Königlichen Akademie der Wissenschaften) pp 844–7
- [3] Galilei G 1638 Dicorsi e dimostrazioni matematiche intorno a due nuove scienze (Lleida: Daniel Adee)
- [4] von Eötvös R, Pekár D and Fekete E 1922 *Ann. Phys.*, *NY* **373**
- [5] Roll P G, Krotkov R and Dicke R H 1964 Ann. Phys., NY 26 442–517
- [6] Braginsky V G and Panov V I 1972 Sov. Phys JETP 34 463-7
- [7] Schlamminger S, Choi K Y, Wagner T A, Gundlach J H and Adelberger E G 2008 Phys. Rev. Lett. 100 041101
- [8] Touboul P et al 2017 Phys. Rev. Lett. 119 231101
- [9] Wiersma D S and Mana G 2021 *Rendiconti Lincei. Sci. Fis. Nat.* **32** 655–63
- [10] Ohanian H C 2012 Am. J. Phys. **80** 1067–72
- [11] Cladé P, Biraben F, Julien L, Nez F and Guellati-Khelifa S 2016 Metrologia 53 A75–82
- [12] Yu C, Zhong W, Estey B, Kwan J, Parker R H and Müller H 2019 Ann. Phys., Lpz. 531 1800346
- [13] Massa E, Sasso C P and Mana G 2020 MAPAN 35 511-9
- [14] Schlamminger S and Haddad D 2019 C. R. Phys. 20 55-63
- [15] Colella R, Overhauser A W and Werner S A 1975 Phys. Rev. Lett. 34 1472–4

Metrologia 59 (2022) 043001 G Mana and S Schlamminger

- [16] Kasevich M and Chu S 1992 Appl. Phys. B 54 321-32
- [17] Cladé P, Guellati-Khélifa S, Schwob C, Nez F, Julien L and Biraben F 2005 Europhys. Lett. 71 730–6
- [18] Littrell K C, Allman B E and Werner S A 1997 Phys. Rev. A 56 1767–80
- [19] Stuhler J, Fattori M, Petelski T and Tino G M 2003 J. Opt. B: Quantum Semiclass. Opt. 5 S75–81
- [20] Wolf P, Blanchet L, Bordé C J, Reynaud S, Salomon C and Cohen-Tannoudji C 2011 Class. Quantum Grav. 28 145017
- [21] Cabiati F 1991 IEEE Trans. Instrum. Meas. 40 110-4
- [22] Kibble B P and Robinson I A 2014 Metrologia 51 S132-9
- [23] Adelberger E G, Gundlach J H, Heckel B R, Hoedl S and Schlamminger S 2009 Prog. Part. Nucl. Phys. 62 102–34
- [24] Wagner T A, Schlamminger S, Gundlach J H and Adelberger E G 2012 Class. Quantum Grav. 29 184002
- [25] Stock M et al 2020 Metrologia 57 07030
- [26] Davidson S and Stock M 2021 Metrologia 58 033002