



ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

Early Biological Response to Poly(ϵ -Caprolactone)/Alumina-Toughened Zirconia Composites
Obtained by 3D Printing for Peri-Implant Application

Original

Early Biological Response to Poly(ϵ -Caprolactone)/Alumina-Toughened Zirconia Composites Obtained by 3D Printing for Peri-Implant Application / Riccardo, Pedraza; Mosca Balma, Alessandro; Roato, Ilaria; Orrico, Clarissa; Genova, Tullio; Baima, Giacomo; Nicolao Berta, Giovanni; Giura, Andrea; Ribotta, Luigi; Duraccio, Donatella; Giulia Faga, Maria; Mussano, Federico. - In: POLYMERS. - ISSN 2073-4360. - 16:(2024). [10.3390/polym16172521]

Availability:

This version is available at: 11696/81759 since: 2024-09-16T12:34:33Z

Publisher:

MDPI

Published

DOI:10.3390/polym16172521

Terms of use:













This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

Article

Early Biological Response to Poly(ϵ -Caprolactone)/Alumina-Toughened Zirconia Composites Obtained by 3D Printing for Peri-Implant Application

Riccardo Pedraza ^{1,2,3,†} , Alessandro Mosca Balma ^{1,†} , Ilaria Roato ¹ , Clarissa Orrico ^{1,3} , Tullio Genova ⁴ , Giacomo Baima ¹ , Giovanni Nicolao Berta ⁵ , Andrea Giura ⁶ , Luigi Ribotta ⁶ , Donatella Duraccio ² , Maria Giulia Faga ²  and Federico Mussano ^{1,*} 

- ¹ Bone and Dental Bioengineering Laboratory, CIR Dental School, Department of Surgical Sciences, University of Turin, 10126 Turin, Italy; riccardo.pedraza@polito.it (R.P.); alessandro.moscabalma@unito.it (A.M.B.); ilaria.roato@unito.it (I.R.); clarissa.orrigo@polito.it (C.O.); giacomo.baima@unito.it (G.B.)
 - ² Institute of Sciences and Technologies for Sustainable Energy and Mobility, National Council of Research, Strada delle Cacce 73, 10135 Turin, Italy; donatella.duraccio@stems.cnr.it (D.D.); mariagiulia.faga@stems.cnr.it (M.G.F.)
 - ³ Department of Mechanical and Aerospace Engineering, Politecnico di Torino, Corso Duca degli Abruzzi 24, 10129 Turin, Italy
 - ⁴ Department of Life Sciences and Systems Biology, University of Turin, Via Accademia Albertina 13, 10123 Turin, Italy; tullio.genova@unito.it
 - ⁵ Department of Clinical and Biological Sciences, University of Turin, Regione Gonzole 10, 10043 Orbassano, Italy; giovanni.bera@unito.it
 - ⁶ Applied Metrology and Engineering Division, Istituto Nazionale di Ricerca Metrologica (INRiM), Strada delle Cacce 91, 10135 Turin, Italy; a.giura@inrim.it (A.G.); l.ribotta@inrim.it (L.R.)
- * Correspondence: federico.mussano@unito.it; Tel.: +39-0116708360
- † These authors contributed equally to this work.



Citation: Pedraza, R.; Mosca Balma, A.; Roato, I.; Orrico, C.; Genova, T.; Baima, G.; Berta, G.N.; Giura, A.; Ribotta, L.; Duraccio, D.; et al. Early Biological Response to Poly(ϵ -Caprolactone)/Alumina-Toughened Zirconia Composites Obtained by 3D Printing for Peri-Implant Application. *Polymers* **2024**, *16*, 2521. <https://doi.org/10.3390/polym16172521>

Academic Editor: Yinghong Chen

Received: 31 July 2024

Revised: 29 August 2024

Accepted: 3 September 2024

Published: 5 September 2024



Copyright: © 2024 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Abstract: The improvement of the mucosal sealing around the implant represents a challenge, one that prompted research into novel materials. To this purpose, a printable poly(ϵ -caprolactone) (PCL)-based composite loaded with alumina-toughened zirconia (ATZ) at increasing rates of 10, 20, and 40 wt.% was prepared, using a solvent casting method with chloroform. Disks were produced by 3D printing; surface roughness, free energy and optical contact angle were measured. Oral fibroblasts (PF) and epithelial cell (SG) tests were utilized to determine the biocompatibility of the materials through cell viability assay and adhesion and spreading evaluations. The highest level of ATZ resulted in an increase in the average roughness (S_a), while the maximum height (S_z) was higher for all composites than that of the unmixed PCL, regardless of their ATZ content. Surface free energy was significantly lower on PCL/ATZ 80/20 and PCL/ATZ 60/40, compared to PCL and PCL/ATZ 90/10. The contact angle was inversely related to the quantity of ATZ in the material. PF grew without variations among the different specimens at 1 and 3 days. After 7 days, PF grew significantly less on PCL/ATZ 60/40 and PCL/ATZ 80/20 compared to unmixed PCL and PCL 90/10. Conversely, ATZ affected and improved the growth of SG. By increasing the filler amount, PF cell adhesion and spreading augmented, while PCL/ATZ 80/20 was the best for SG adhesion. Overall, PCL/ATZ 80/20 emerged as the best composite for both cell types; hence, it is a promising candidate for the manufacture of custom made transmucosal dental implant components.

Keywords: poly(ϵ -caprolactone); alumina-toughened zirconia; solvent casting; 3D printing; cell viability; early cell response; oral keratinocytes; oral fibroblasts

1. Introduction

Over the last few decades, dental implants have emerged as the preferred treatment option for rehabilitating edentulism, since they represent a reliable and enduring solution which can be used to reestablish the aesthetics and function of natural teeth, even

with immediate loading protocols [1]. Also, the neoformation of bone tissue in continuity with implant fixtures—known as osseointegration—prevents post-extractive jaw bone resorption [2]. Most dental implant systems on the market are characterized by roughened titanium surfaces, which have been shown in the literature to promote osseointegration [3]. These surfaces have been pointed out, albeit not unanimously [4], as favorable environments for biofilm accumulation, and hence the onset of peri-implantitis [5], i.e., the progressive bone loss around the fixture associated with the microbial colonization.

Owing to the high prevalence of peri-implantitis [6], researchers oriented their studies to the prevention of this disease, using a two-fold approach by introducing antibacterial surfaces on the intraosseous fixtures [7] and enhancing the adhesion of the peri-implant tissues [8]. Attaining a strong mucosal seal of the abutments, i.e., the transmucosal components of the implants, has become, therefore, paramount in order to prevent implant failure [9]. Promising results were achieved, for instance, with a Poly(dopamine)-modified alkali-heat-titanium surface enriched with hydroxyapatite and carboxymethyl chitosan, as it could enhance human gingival fibroblast adhesion, spread and proliferation, also ensuring antibacterial activity [10]. Even anodization of titanium has seemed beneficial for reducing the colonization of representative bacterial strains [11], while having remarkable effects on fibroblast proliferation only at the nanoscale [12].

These approaches tackled only the connective component, neglecting the epithelium, which is the outermost barrier between the organism and the environment. Yet the peri-implant epithelium, usually named the “long junctional epithelium” [13], results from the downward migration of the oral epithelium parallel to the abutment, which is only stopped by the connective tissue with its circumferential collagen fibers surrounding the abutment [14]. Lacking perpendicular fibers, the peri-implant connective tissue is not as tightly adherent as the periodontal connective tissue, and thus it may allow the junctional epithelium to penetrate too deeply, resulting ultimately in bone loss [15]. A key achievement of the research would be to develop a bio-interface capable of promoting an improved biomimicry of the normal periodontium.

Such an ambitious goal would entail the re-formation of an internal basal lamina between the junctional epithelium and the abutment, capable of limiting the former’s downward migration. To this purpose, among dental implant biomaterials [14], alumina-toughened zirconia (ATZ) appears as a promising option based on its peculiar capacity of orienting laminin 332, which is indispensable for guiding proper epithelial cell adhesion [16]. Besides this remarkable result achieved *in vitro*, ATZ could also outperform titanium when used as an implant material in a minipig model [17]. As a massive ceramic material, however, ATZ is hindered by its poor mechanical properties [18]. For this reason, combining the bioactivity of this material with polymers which are endowed with high tenacity although almost biologically inert, like UHMWPE, has recently become a topic of interest [19].

Among the numerous polymers available for regenerative medicine, the Poly(ϵ -caprolactone) (PCL) holds particular interest, being a biocompatible, synthetic, aliphatic polyester that, besides a widely diffused application for intraosseous scaffolds [20], has more recently revealed a remarkable suitability in periodontal ligament regeneration [21]. After proper manipulation [22] PCL could indeed guide fibroblast adhesion. This feature, together with its use as epidermal equivalent [23], and its versatility in the preparation of bioresorbable coatings [24], renders PCL the candidate of choice for preparing fibro-mucosal interfaces along the abutments.

Recently, our group obtained PCL-ATZ composites through solvent casting with better results than obtained with the dry mixing technique [25], paving the way to the easy tuning of the mechanical, and possibly biological, properties of these biomaterials. In the present study, the authors characterized *in vitro* the biological responses of three different ATZ/PCL compounds, with respect to oral fibroblasts and keratinocytes, in established cell models of gingival tissue. The purpose of this research was to assess whether the cell response could be modulated by simply varying the concentration of ATZ within the

PCL matrix, with the aim of selecting a formulation possibly useful for the fabrication of functionalized implant abutments to ameliorate difficulties in the peri-implant mucosal seal.

2. Materials and Methods

2.1. Sample Preparation

An ester-terminated polycaprolactone (CAS-n 24980-41-4) matrix (CELLINK PCL TP-60505, Bico Group, Gothenburg, Sweden) was charged with alumina-toughened zirconia (ATZ, made of 20 wt.% Al_2O_3 and 80 wt.% 3Y-TZP composed of 3 mol% yttria-stabilized zirconia, Tosoh Bioscience, Tokyo, Japan) through solvent casting with chloroform (CHCl_3 , CARLO ERBA Reagents s.r.l., Cornaredo, Italy), as previously described [25]. Three PCL/ATZ composite materials were prepared: PCL/ATZ 90/10, PCL/ATZ 80/20, and PCL/ATZ 60/40. Unmixed PCL was considered as a control.

Three-dimensional planar samples were printed using a thermoplastic pneumatic printhead with a BIO X 3D bioprinter (CELLINK Bico Group, Gothenburg, Sweden). A standard square base geometry of 15 mm \times 15 mm with a thickness of 0.65 mm and a 100% infill was printed directly on the glass surface of a Petri dish in order to obtain the smoothest interaction surface possible for each compound. These samples were used for surface roughness and contact-angle measurements. Cylindrical discs were obtained by cutting the previously described square 3D printed geometry with a 6mm biopsy punch, and were used for protein adsorption, cell adhesion, cell spreading, and cell viability tests, as well as SEM analysis. The print-bed temperature was set at 30 °C, with a clean chamber fan kept on. Printhead temperature was 115 °C for unmixed PCL, 125 °C for PCL/ATZ 90/10, 135 °C for PCL/ATZ 80/20, and 145 °C for PCL/ATZ 60/40. A nozzle 0.4 mm in diameter was employed; the imposed pressure was 190 kPa, and the printing speed was set at 2 mm/s.

2.2. Microscopy

Microstructure was studied by means of a Scanning Electron Microscope (Phenom XL G2 Desktop SEM, Thermo Fisher Scientific, Waltham, MA, USA). Before examination, the samples were: (a) washed in distilled water; (b) rinsed thoroughly in 70% ethanol–water solution; (c) cleaned ultrasonically in absolute ethanol for 20 min; (d) air dried under a chemical hood; and (e) coated with a thin conductive layer of gold. The instrument settings adopted for micrographs and analyses were 10 kV of voltage v in a MAP configuration with a Back Scatter Detector (BSD).

2.3. Roughness

Surface roughness was calculated according to ISO 21920 (for profile R parameters) and ISO 25178 (for areal S parameters) on unmixed PCL, PCL/ATZ 90/10, PCL/ATZ 80/20, and PCL/ATZ 60/40 samples; the analysis was carried out by both contact-based and optical profilometers. The contact-based test was made with a stylus profilometer (Form Talysurf PGI Novus S 10, Taylor Hobson Limited, Leicester, UK) equipped with a precision ceramic ball, which is calibrated by an interferometric setup to be traceable to the International Systems of Unit (SI). During the measurements, a cut-off of 0.08 mm was employed, and a total length of 8 mm was set. With the contact technique, roughness parameters Arithmetical Mean Height (R_a), Root Mean Square Height (R_q), Maximum Height (R_z), Skewness (R_{sk}) and Kurtosis (R_{ku}) were calculated.

The second technique utilized the optical profilometer (Sensofar Plμ 2300, Barcelona, Spain) in confocal mode to perform the topography measurements. As before, the cut-off was set at 0.08 mm and the surface size was 254.64 μm \times 190.90 μm , since a Nikon LU Plan Fluor 50 \times /0.80 objective was used. With this second technique, Arithmetical Mean Height (S_a), Root Mean Square Height (S_q), Maximum Height (S_z), Skewness (S_{sk}) and Kurtosis (S_{ku}) surface parameters were measured; so not to ruin the specimen, optical measurements were performed before tactile measurements. For each sample, 5 optical measurements and

3 stylus measurements were carried out. The topographies and the profiles were processed by using MountainsMap Premium 10 by Digital Surf (Besançon, France).

2.4. Contact Angle and Surface Free Energy Evaluation

Surface wettability was assessed using a Biolin Scientific Theta Lite Optical Tensiometer (Stockholm, Sweden) using double-distilled water (dH₂O) and diiodomethane (CH₂I₂). The contact angle was evaluated by the sessile drop method. For each liquid drop (1 µL in volume) dispensed, an image of the drop on the sample was acquired with the integrated high-resolution camera. The drop profiles were extracted and fitted with integrated software. At the liquid–solid interface, contact angles between fitted function and base line were calculated. For each sample and liquid probe, the contact-angle measurement was repeated five times on different areas.

The Owens–Wendt–Rabel–Kaelble (OWRK) method was adopted to calculate a value of surface free energy, following the method proposed by Waldner, C. et al. [26]. Total (γ), polar (γ^P), and dispersive (γ^D) components were calculated by simple linear regression. Properties of dH₂O and CH₂I₂ were taken as standard constants to perform the interpolation, as reported in Table 1 [27].

Table 1. Standard parameters of water and di-iodomethane.

Liquid	γ [mN/m]	γ^P [mN/m]	γ^D [mN/m]
Water	72.8	43.7	29.1
Di-iodomethane	50	2.6	47.4

2.5. Protein Adsorption

To quantify the amount of protein adsorbed onto the samples, as described before [28], a 5% solution of Bovine Serum Albumin (BSA) in Phosphate Buffered Saline (PBS) was prepared and used to cover all the compounds. Specimens were incubated at 37 °C for 20 min, and then were washed twice with PBS. The total adsorbed protein amount was determined by means of elution with Tris Triton buffer (10 mM Tris (pH 7.4), 100 mM NaCl, 1 mM EDTA, 1mM EGTA, 1% Triton X-100, 10% Glycerol and 0.1% SDS) for 10 min, with the result subsequently quantified through Pierce™ BCA Protein Assay Kit (Life Technologies, Carlsbad, CA, USA) according to the manufacturer’s instructions.

2.6. Cell Experiments

2.6.1. Cell Culture

Primary Palatal Fibroblasts (PF) and human gingival epithelioid cell line (SG) [29] were utilized to characterize the biological response in vitro, as induced by the different printed samples. PF were cultured in Alpha-MEM (Life Technologies, Milano, Italy), 10% fetal bovine serum (FBS, Life Technologies, Milan, Italy), and 5% penicillin (100 U/mL)-streptomycin (100 µg/mL) at 37 °C and with a 5% CO₂ atmosphere. SG cells were maintained in an RPMI-1640 medium (Euroclone, Pero, Italy) with 10% FBS, 100 U/mL penicillin, and 100 µg/mL streptomycin.

2.6.2. Cell Adhesion and Cell Spreading

Cells were seeded at 4000 cells/well on the samples with different percentages of filler and incubated for 40 min, then fixed by 4% paraformaldehyde solution, and washed with PBS; cell nuclei were stained by 1 µM DAPI (Merck, Darmstadt, Germany) for 15’ at 37 °C.

For the cell spreading evaluation, cells were maintained in culture for 24 h, then fixed and stained with Phalloidin (Cell Signaling technology, Danvers, MA, USA) for cytoskeleton and with DAPI for cell nuclei.

Images were acquired using a Nikon Eclipse Ti-E microscope with Nikon Plan 40×/0.75 and Nikon Plan 10×/0.10 objectives. Cell nuclei were counted using the ‘Analyze particles’ tool of ImageJ software (Version 2.14.0/1.54f, ImageJ, U. S. National Institutes of Health,

Bethesda, MD, USA). For cell spreading analysis, 4 images were acquired at a “higher magnification” for each sample type in triplicate, and were firstly processed with the cellpose [30,31] cyto3 segmentation algorithm, and then pre-trained with other similar images in a controlled learning process, to obtain the contours of the single cell on all the acquired fields. Then, 12 different shape descriptors were measured for each detected cell with the Set Measurements function implemented in Fiji/ImageJ (i.e., area, perimeter, best fitting ellipse (BFE) major axis, BFE minor axis, BFE aspect ratio, BFE angle, circularity, roundness, solidity, Feret’s diameter, Feret’s angle, and minimum caliper diameter). From these 12 descriptors, 6 were selected as relevant parameters for cell morphology evaluation (area, perimeter, BFE aspect ratio, BFE angle, circularity, and roundness). Data were analyzed and plotted in MATLAB (MATLAB R2024a; The MathWorks, Inc., Natick, MA, USA).

2.6.3. Cell Viability

PF and SG were plated on the different samples at a density of 10,000 cells/well in 96-well culture plates, in their cell culture media. Cell viability was assessed after 1, 3, and 7 days of in vitro culture through use of the Cell Titer GLO kit (Promega, Milan, Italy) according to the manufacturer’s protocol. Cell viability was expressed as relative light unit (RLU).

2.7. Statistical Analysis

Statistical analysis was performed through STATA software (version 18.0; StataCorp, College Station, TX, USA). One-way ANOVA was performed to evaluate differences among group variances at different time steps, and a Bonferroni post hoc corrective coefficient was adopted to find which groups differed in a statistically relevant way. Repeated *t*-tests were used to perform statistical analyses on shape descriptors for cell morphology evaluation. An α significance level of 0.05 was utilized [32,33].

3. Results

3.1. SEM Morphological Surface Analysis

Figure 1a shows the unmixed PCL surface, while PCL/ATZ 90/10, PCL/ATZ 80/20, and PCL/ATZ 60/40 surfaces are represented in Figure 1b, 1c, and 1d, respectively. In composites, the augmenting presence of ATZ filler is indicated by increasing numbers of white dots. Although all the samples were printed on the same glass support to obtain the smoothest and the most uniform possible surface, the ATZ distribution became more homogeneous proportionally to the increase of the amount of filler. Indeed, some residual stripes of ATZ aligned to the printing direction, likely due to the passage of the print head during the fused deposition process (Figure 1b,c), this disappeared when the highest ATZ percentage was used.

3.2. Roughness Analysis

PCL/ATZ 60/40 samples reached the highest roughness values, as reported from the averages R_a and S_a and the root-mean-square values R_q and S_q (Tables 2 and 3, respectively). Skewness parameters (S_{sk}) were approximately 1.5 for both unmixed PCL and PCL/ATZ 60/40, indicating an asymmetrical distribution of peaks and valleys in which more peaks are present than valleys. A large standard deviation was observed for these values due to the large exponent used for their calculation; indeed, parameters S_{sk} and S_{ku} are sensitive to the presence of spikes as well as narrow and deep valleys. However, the measurements were repeated to find whether there was a positive or negative tendency in the mean values. PCL/ATZ 80/20 showed values of Skewness that were near to 0, representing a homogeneous distribution of peaks and valleys which are very sharp in their profile, as highlighted by values of Kurtosis (S_{ku}) greater than 3. All the samples were characterized by a leptokurtic distribution of valleys or peaks ($S_{ku} > 3$). All composites showed higher Maximum Height (S_z) values than unmixed PCL, regardless of the ATZ amount.




Figure 1. Scanning electron micrographs of materials: (a) unmixed PCL, (b) PCL/ATZ 90/10, (c) PCL/ATZ 80/20, and (d) PCL/ATZ 60/40.

Table 2. Profile roughness R parameters for the different samples measured with the stylus profilometer.

Sample	R_a [μm]	R_q [μm]	R_{sk}	R_{ku}	R_z [μm]
PCL	0.103 ± 0.019	0.173 ± 0.035	3.273 ± 0.525	40.284 ± 15.079	0.613 ± 0.107
PCL/ATZ 90/10	0.122 ± 0.016	0.234 ± 0.102	-1.597 ± 2.772	85.616 ± 43.422	0.711 ± 0.063
PCL/ATZ 80/20	0.106 ± 0.002	0.143 ± 0.007	-0.644 ± 0.594	10.712 ± 9.004	0.641 ± 0.010
PCL/ATZ 60/40	0.207 ± 0.001	0.306 ± 0.002	-1.003 ± 0.252	11.909 ± 2.799	1.296 ± 0.015

Table 3. Areal texture S parameters for the different samples measured with the optical profilometer.

Sample	S_a [μm]	S_q [μm]	S_{sk}	S_{ku}	S_z [μm]
PCL	0.122 ± 0.021	0.238 ± 0.066	1.825 ± 2.553	27.887 ± 22.966	4.233 ± 2.141
PCL/ATZ 90/10	0.135 ± 0.007	0.227 ± 0.016	-1.049 ± 1.461	91.515 ± 91.120	13.005 ± 6.984
PCL/ATZ 80/20	0.062 ± 0.004	0.102 ± 0.017	-0.123 ± 12.575	1400.300 ± 1223.800	13.861 ± 5.640
PCL/ATZ 60/40	0.155 ± 0.035	0.248 ± 0.057	1.515 ± 5.255	163.070 ± 217.170	14.631 ± 4.138

3.3. Wettability and Surface Free Energy Evaluation

The responses of the materials to both hydrophilic and lipophilic environments were assessed by returning to the standard optical contact angle (OCA) measurement with a polar solvent and an apolar solvent. The measured dH_2O contact angles for all materials remained stable at approximately 67° , with no relevant differences found among the samples (Figure 2a). The only specimen that demonstrated a slight, but not significant, difference was PCL/ATZ 80/20 (Figure 2a). In a lipophilic environment with CH_2I_2 , unmixed PCL showed the highest contact-angle value (41°), while all composite materials presented a decreased contact angle, which was 36° in PCL/ATZ 90/10 and 23° in both PCL/ATZ 80/20 and PCL/ATZ 60/40 (Figure 2b). The only statistically significant difference was detected between PCL and PCL/ATZ 80/20 ($p < 0.05$).




Figure 2. Graphs representing surface contact angle calculated in a hydrophilic environment (dH_2O , (a)) and in a lipophilic environment (CH_2I_2 , (b)) (* for $p < 0.05$).

The SFE calculation showed higher values of total surface energy (γ) in the presence of the polymer alone (4.15) or with lesser ATZ filler (PCL/ATZ 90/10), and lower values of γ for PCL/ATZ 80/20 (4.02) and PCL/ATZ 60/40 (4.09). This behavior was also seen for the polar component of surface energy, although there were no significant variations in the values of dispersive surface energy, as shown in Table 4.

Table 4. Total, polar, and dispersive surface energy of samples, calculated with the OWRK method.

Sample	Surface Energy: Total [mN/m]	Surface Energy: Polar [mN/m]	Surface Energy: Dispersive [mN/m]
PCL	4.15	1.77	2.38
PCL/ATZ 90/10	4.14	1.72	2.43
PCL/ATZ 80/20	4.02	1.48	2.54
PCL/ATZ 60/40	4.09	1.56	2.53

3.4. Protein Adsorption

The protein adsorption graph (Figure 3) shows that the amount of BSA protein adsorbed on the sample surfaces tended to increase with the ceramic filler percentage. Unmixed PCL displayed the lowest value (~ 15 mg/mL), compared to composites. In particular, PCL/ATZ 60/40 showed the highest value (~ 45 mg/mL) among all tested conditions with results significantly different from those of the pure polymer ($p < 0.05$). While PCL/ATZ 80/20 and PCL/ATZ 90/10 were similar in mean value, only the former differed from PCL in a statistically significant way ($p < 0.05$).

3.5. Cell Experiments

3.5.1. Cell Adhesion and Cell Spreading

All of the samples allowed cell adhesion. Regarding PF (Figure 4a), notably, on composites, the number of adherent cells was augmented proportionally to the amount of filler. The PCL/ATZ 90/10 supported the lowest number of cellular nuclei compared to PCL/ATZ 80/20 and PCL/ATZ 60/40 (respectively, $p = 0.03$ and $p = 0.008$), which showed increasing numbers of adherent cells. No composite, however, outperformed the PCL in a statistically significant way.




Figure 3. Graph representing the amount of BSA protein (mg/mL) adsorbed on specimen surfaces ((*) for $p < 0.05$).




Figure 4. Graph representing PF (a) and SG (b) cell adhesion, as evaluated after 20 min of incubation ((*) for $p < 0.05$ and (**) for $p < 0.01$).

The incorporation of ATZ particles in the polymeric matrix enhanced, instead, the adhesive capabilities of the SG cells (Figure 4b). This result was particularly evident for PCL/ATZ 80/20, in which the highest number of adherent cells was achieved overall in a statistically significant way ($p < 0.05$).

The cell spreading analysis using the BFE angle in segmented cells is represented below in a polar histogram plot (Figure 5). The PF spreading, in this case, lost its partial orientation as the ATZ amount was raised inside the PCL matrix. This type of behavior was not confirmed with the SG cells, possibly due to their more rounded shape and their random distributions on surfaces.

Area and perimeter values measured on PF images (Figure 6) were higher for the compounds with ATZ than the unmixed PCL samples. PCL/ATZ 80/20 and PCL/ATZ 60/40 presented the (significant) highest values overall for both the descriptors ($p < 0.05$). Consistent with this, the aspect ratio values were also significantly lower for the pure polymer compared to the others. The complex descriptors of circularity and roundness point out a statistically relevant difference between the PCL/ATZ 90/10 values and the unmixed PCL ($p = 0.004$ and $p = 0.017$). Overall, the total area occupied by the cell accorded with the above-described cell adhesion test (Figure 4a).




Figure 5. Polar histogram plot of BFE angles measured on each segmented PF cell for all the compounds (in blue are represented the polar bins of the distributions, while in red are highlighted the mean angle values).

PCL/ATZ 80/20 had the (significantly) smallest dimensions of cell spreading for SG in terms of area and perimeter values. The aspect ratio, circularity, and roundness of the cell were, respectively, significantly lower for the first, and higher for the second and third in the PCL samples not filled with ATZ ($p < 0.05$ and $p < 0.01$). The sum of all the area values reported was according to the adhesion test (Figure 4b).

3.5.2. Cell Viability

At 1 and 3 days, the PF cells grew without significant variations among the different specimens. After 7 days, PF cells proliferated significantly less on PCL/ATZ 60/40 ($p < 0.01$) and PCL/ATZ 80/20 ($p < 0.05$), compared to unmixed PCL and PCL/ATZ 90/10 (Figure 7a).

SG viability showed absolute values lower than those for PF cells. At day 3, SG cells grew significantly better on PCL/ATZ 80/20 compared to the composite materials ($p < 0.05$). At day 7, the growth of SG cells on composite materials was equivalent and significantly higher compared to the unmixed PCL ($p < 0.01$, Figure 7b), suggesting that the filler positively affected the proliferation of SG cells.




Figure 6. Immunofluorescence microscopy images representing PF cells on (a) unmixed PCL, (b) PCL/ATZ 90/10, (c) PCL/ATZ 80/20, and (d) PCL/ATZ 60/40, and SG cells on (e) unmixed PCL, (f) PCL/ATZ 90/10, (g) PCL/ATZ 80/20, and (h) PCL/ATZ 60/40, after 24 h of incubation (the nuclei of cells are represented in blue, while the cytoskeletons were stained with green).




Figure 7. Graph representing PF (a) and SG (b) cell viability at 1, 3, and 7 days (*) for $p < 0.05$ and (**) for $p < 0.01$.

4. Discussion

The development of an interface material to enhance the mucosal attachment to the implant abutments could limit the probability of peri-implantitis. This work had the purpose of implementing an innovative material, one based on a PCL matrix filled with different amounts of ATZ (10, 20, and 40 wt.%), by using a previously described solvent casting method [25]. After a chemical and physical characterization of the surfaces, these composites were studied as for the early biological response elicited using the two cell types (PF and SG) representative of the mucosal tissue [29,34].

All of the compounds and the unmixed PCL were 3D-printed in regular shapes on the same glass substrate to achieve the smoothest possible surface. This aimed at limiting the variability the additive manufacturing process could have on surface roughness, which is known to play a key role, and may overshadow other features of the bulk material, when characterizing its interface [35–38]. Nevertheless, SEM micrographs of these surfaces showed the formation of residual stripes of filler that were visible with PCL/ATZ 90/10 and PCL/ATZ 80/20, and oriented parallel to the printing direction. This phenomenon could be due to the swelling ratio of the polymer during the extrusion and cooling process [39,40], which is responsible of changes in volume and viscosity, and gives the filler the tendency to aggregate at the interface between strands. As described by Bellini [36] and Shofner et al. [41,42], the value of the swelling ratio depends on both the material properties and the geometry of the extrusion nozzle. Indeed, inelastic fillers such as ceramics could reduce die swelling, as was also observed here for PCL/ATZ 60/40, with which these stripes were not present, due to the lower ratio of polymer to filler, which allowed a more uniform surface to be obtained.

In this work, the S_z values were augmented proportionally with the composites, consistently with the fact that PCL was expected to be the smoothest surface. Also, PCL/ATZ 60/40 samples showed the highest roughness values (i.e., R_a , S_a , R_q , and S_q values). The introduction of ATZ in the polymer matrix affected the S_{sk} of the compounds, ranging from the negative value of PCL/ATZ 90/10 (−1.049) to the positive value of PCL/ATZ 60/40 (1.515), and reaching a value close to 0 for PCL/ATZ 80/20. This means that an asymmetrical distribution of peaks and valleys was present on the two composites with the lowest and the highest amount of ATZ; these had, respectively, a prevalence of valleys and peaks, while a homogeneous distribution thereof characterized PCL/ATZ 80/20. Less predictable outcomes in terms of roughness parameters were achieved for PCL/ATZ 80/20. This compound showed S_a values that were half of those from the other samples, and it reached the highest values by far (almost ten times the other specimens) of Kurtosis ($S_{ku} = 1400.300 \pm 1223.800$), which describes very sharp profiles.

According to the wettability assay, all the materials showed a mild hydrophilicity (dH_2O CA around 67°), with a general preference indicated for lipophilic environments,

which is consistent with similar polymeric materials [43]. Furthermore, the filler appeared to amplify this behavior, although a statistically significant difference was achieved only between PCL ($CA = 41^\circ$) and PCL/ATZ 80/20 ($CA = 23^\circ$). Overall, this trend was reflected by the moderately higher values of total surface energy (γ) of the PCL (4.15 mN/m) and PCL/ATZ 90/10 (4.14 mN/m), compared to the other two compounds richer in ATZ. These data are in accordance with the high hydrophobicity of ATZ reported by Carvalho et al. [44], as its addition to PCL could enhance the dispersive component of SFE, albeit not significantly. A higher quantity of ATZ also facilitated protein adsorption on the samples' surfaces, with a three-fold increase found when PCL/ATZ 60/40 (~45 mg/mL) was compared to PCL (~15 mg/mL). Since all the compounds were made of the same PCL matrix, it is reasonable to assume that the amounts of ATZ particles available on the surfaces provided preferable sites for protein linkage [16], contrary to the findings of Yang et al. [45], in which BSA adsorption on nano-roughened titanium was found to scale linearly with S_q .

Analyzing the early biological response, PF followed the trend of S_{sk} in their adhesion pattern, which is an unprecedented observation, to the authors' knowledge. On the contrary, neither R_{ku} nor S_{ku} seemed correlated with the PF's responses to the surfaces tested, which is different from the findings of Frias Martinez et al. [46], in which the samples with higher values of R_{ku} favored the adhesion of human gingival fibroblasts. The importance of Skewness and Kurtosis in orienting cell differentiation was indeed elucidated a decade ago [47] in human mesenchymal cells and osteoblasts. The morphological analysis of cell spreading showed that as the amount of ATZ increased, PF: (a) lost progressively their partial orientation parallel to the printing moving direction, (b) exhibited larger areas, and (c) developed a higher number of filopodia ramifications on the cell borders, compared to PCL. As also confirmed by the values for cell aspect ratio, PCL/ATZ 90/10 and PCL/ATZ 80/20 emerged as the preferable surfaces for PF spreading.

Quite interesting was the correlation between the specific roughness parameters of PCL/ATZ 80/20 and the early cell behavior of SG that adhered the most on this surface in a statistically significant way (Figure 4b). It is conceivable that the topographic pattern of PCL/ATZ 80/20, endowed with low S_a , high S_{ku} , and balanced S_{sk} , resulted in the most suitable environment for early adhesion of SG. Unfortunately, the adhesion properties of SG seeded on bio-interfaces, and more generally those of oral keratinocytes, have been poorly studied so far, and the study has been limited to a few materials [48]. Analyzing the morphology of SG, a relevant difference in terms of aspect ratio, circularity, and roundness emerged, owing to their rounded and less spread shape. The surface of unmixed PCL was the case where SG were more rounded in shape. On the contrary, the ATZ particles on the surface appeared to be favorable sites for cell adhesion, although suffering a cell tolerance cap at PCL/ATZ 80/20, since PCL/ATZ 60/40 showed decreasing adhesion-based features. Noteworthy, the favorable effect described on adhesion on PCL/ATZ 80/20 was not obliterated immediately, as this surface still outperformed the other ones, after 3 days, in terms of cell proliferation. At 7 days, instead, all the compounds were remarkably superior to PCL as to cell viability, consistent with Saberian et al. [49]. Unlike SG, PF grew without significant variations among the specimens at 1 and 3 days, but they proliferated significantly less on PCL/ATZ 60/40 ($p < 0.01$) and PCL/ATZ 80/20 ($p < 0.05$), compared to PCL, at day 7. Such a late decrease in the proliferation of PF is unlikely to be due to a lack of biocompatibility, but it is likely ascribable to the presence of differentiative cues. Under this perspective, it is compelling that PCL/ATZ 80/20 could promote the highest number of focal adhesions per cell in PF (Figure S1), resulting the best condition for facilitating early cell adhesion in both the cell models used. Finally, the remarkable divergence concerning the proliferation rates of PF and SG should not be matter of surprise, since cells, depending on their type, may behave quite differently on the same surface [50]. One theoretical explanation of these dissimilar patterns may rely upon the machinery used by PF and SG for interacting with their substrates. While, generally, fibroblasts adhere on a given surface through the formation of classic focal adhesion complexes, epithelial cells lay

on and anchor to a basal lamina, requiring specific proteins like laminins, as reported by Riivari et al. [51].

In summary, PCL, when mixed with ATZ, is endowed with certain non-negligible characteristics, such as improved protein adsorption and cell adhesion. Also, the PCL/ATZ compounds could affect cell spreading and the viability of both PF and SG. Specifically, PCL/ATZ 80/20 outperformed all of the other conditions when used as interface for SG and behaved in a manner sufficient to sustain PF viability and adhesion. Hence, such a compound could become a suitable candidate for testing in further studies aiming to functionalize the transmucosal component of dental implants and to improve the mucosal seal, and thus reduce the risk of peri-implantitis.

Supplementary Materials: The following supporting information can be downloaded at: <https://www.mdpi.com/article/10.3390/polym16172521/s1>, Figure S1. Immunofluorescence microscopy images representing the following: (a) PF cell on unmixed PCL surface; (b) PF cell on PCL/ATZ 80/20 surface (the nuclei of cells are represented in blue, while the cytoskeletons were stained with a green color; red arrows highlight the focal adhesion spots).

Author Contributions: Conceptualization, A.M.B., R.P. and F.M.; methodology, A.M.B., R.P., I.R., G.B. and C.O.; software, A.M.B., R.P. and I.R.; validation, I.R., F.M., D.D., M.G.F. and T.G.; formal analysis, A.M.B., R.P., I.R., A.G., L.R. and T.G.; investigation, I.R., F.M., D.D. and M.G.F.; resources, F.M.; data curation, A.M.B., R.P., G.B., I.R., A.G. and L.R.; writing—original draft preparation, A.M.B., R.P., G.N.B., I.R. and F.M.; writing—review and editing, A.M.B., R.P., M.G.F., I.R. and F.M.; visualization, I.R. and F.M.; supervision, F.M.; project administration, D.D., G.N.B., M.G.F. and F.M.; funding acquisition, I.R. and F.M. All authors have read and agreed to the published version of the manuscript.

Funding: Project Title—“Multiscale modelling/characterization and fabrication of nanocomposite ceramics with improved toughness (CONCERTO)”.

Institutional Review Board Statement: Not applicable.

Data Availability Statement: The original contributions presented in this study are included in the article. Further inquiries can be directed to the corresponding author.

Acknowledgments: We wish to thank Paola Iacomussi and Marco Pisani heartily for the kind support and suggestions.

Conflicts of Interest: The authors declare no conflicts of interest.

References

1. Garcia-Sanchez, R.; Dopico, J.; Kalemaj, Z.; Buti, J.; Pardo Zamora, G.; Mardas, N. Comparison of clinical outcomes of immediate versus delayed placement of dental implants: A systematic review and meta-analysis. *Clin. Oral Implant. Res.* **2022**, *33*, 231–277. [\[CrossRef\]](#) [\[PubMed\]](#)
2. Pinheiro, F.A.L.; Mourão, C.F.d.A.B.; Diniz, V.S.; Silva, P.C.; Meirelles, L.; Santos, E.; Schanaider, A. In-vivo bone response to titanium screw implants anodized in sodium sulfate. *Acta Cirúrgica Bras.* **2014**, *29*, 376–382. [\[CrossRef\]](#) [\[PubMed\]](#)
3. Duraccio, D.; Mussano, F.; Faga, M.G. Biomaterials for dental implants: Current and future trends. *J. Mater. Sci.* **2015**, *50*, 4779–4812. [\[CrossRef\]](#)
4. Stavropoulos, A.; Bertl, K.; Winning, L.; Polyzois, I. What is the influence of implant surface characteristics and/or implant material on the incidence and progression of peri-implantitis? A systematic literature review. *Clin. Oral Implant. Res.* **2021**, *32*, 203–229. [\[CrossRef\]](#) [\[PubMed\]](#)
5. Raes, M.; D’hondt, R.; Teughels, W.; Coucke, W.; Quirynen, M. A 5-year randomized clinical trial comparing minimally with moderately rough implants in patients with severe periodontitis. *J. Clin. Periodontol.* **2018**, *45*, 711–720. [\[CrossRef\]](#)
6. Fu, J.H.; Wang, H.L. Breaking the wave of peri-implantitis. *Periodontology 2000* **2020**, *84*, 145–160. [\[CrossRef\]](#)
7. Mandracci, P.; Mussano, F.; Rivolo, P.; Carossa, S. Surface treatments and functional coatings for biocompatibility improvement and bacterial adhesion reduction in dental implantology. *Coatings* **2016**, *6*, 7. [\[CrossRef\]](#)
8. Laleman, I.; Lambert, F. Implant connection and abutment selection as a predisposing and/or precipitating factor for peri-implant diseases: A review. *Clin. Implant Dent. Relat. Res.* **2023**, *25*, 723–733. [\[CrossRef\]](#)
9. Gibbs, S.; Roffel, S.; Meyer, M.; Gasser, A. Biology of soft tissue repair: Gingival epithelium in wound healing and attachment to the tooth and abutment surface. *Eur. Cells Mater.* **2019**, *38*, 63–78. [\[CrossRef\]](#)

10. Yang, M.; Jiang, P.; Ge, Y.; Lan, F.; Zhou, X.; He, J.; Wu, Y. Dopamine self-polymerized along with hydroxyapatite onto the preactivated titanium percutaneous implants surface to promote human gingival fibroblast behavior and antimicrobial activity for biological sealing. *J. Biomater. Appl.* **2018**, *32*, 1071–1082. [\[CrossRef\]](#)
11. Genova, T.; Chinigò, G.; Munaron, L.; Rivolo, P.; Luganini, A.; Gribaudo, G.; Cavagnetto, D.; Mandracchi, P.; Mussano, F. Bacterial and Cellular Response to Yellow-Shaded Surface Modifications for Dental Implant Abutments. *Biomolecules* **2022**, *12*, 1718. [\[CrossRef\]](#)
12. Deng, Z.; Yu, L.; Kuang, Y.; Zhou, Z.; Li, X. Highly Ordered Nanotube-Like Microstructure on Titanium Dental Implant Surface Fabricated via Anodization Enhanced Cell Adhesion and Migration of Human Gingival Fibroblasts. *Int. J. Nanomed.* **2024**, *19*, 2469–2485. [\[CrossRef\]](#) [\[PubMed\]](#)
13. Parpaiola, A.; Cecchinato, D.; Toia, M.; Bressan, E.; Speroni, S.; Lindhe, J. Dimensions of the healthy gingiva and peri-implant mucosa. *Clin. Oral Implant. Res.* **2015**, *26*, 657–662. [\[CrossRef\]](#)
14. Rompen, E.; Domken, O.; Degidi, M.; Farias Pontes, A.E.; Piattelli, A. The effect of material characteristics, of surface topography and of implant components and connections on soft tissue integration: A literature review. *Clin. Oral Implant. Res.* **2006**, *17*, 55. [\[CrossRef\]](#)
15. Schierano, G.; Ramieri, G.; Cortese, M.; Aimetti, M.; Preti, G. Organization of the connective tissue barrier around long-term loaded implant abutments in man. *Clin. Oral Implant. Res.* **2002**, *13*, 460–464. [\[CrossRef\]](#)
16. Vallée, A.; Faga, M.; Mussano, F.; Catalano, F.; Tolosano, E.; Carossa, S.; Altruda, F.; Martra, G. Alumina–zirconia composites functionalized with laminin-1 and laminin-5 for dentistry: Effect of protein adsorption on cellular response. *Colloids Surf. B: Biointerfaces* **2014**, *114*, 284–293. [\[CrossRef\]](#) [\[PubMed\]](#)
17. Schierano, G.; Mussano, F.; Faga, M.G.; Menicucci, G.; Manzella, C.; Sabione, C.; Genova, T.; Degerfeld, M.M.v.; Peirone, B.; Cassenti, A. An alumina toughened zirconia composite for dental implant application: In vivo animal results. *BioMed Res. Int.* **2015**, *2015*, 157360. [\[CrossRef\]](#)
18. Maji, A.; Choubey, G. Microstructure and mechanical properties of alumina toughened zirconia (ATZ). *Mater. Today: Proc.* **2018**, *5*, 7457–7465. [\[CrossRef\]](#)
19. Duraccio, D.; Strongone, V.; Faga, M.; Auriemma, F.; Mussano, F.; Genova, T.; Malucelli, G. The role of different dry-mixing techniques on the mechanical and biological behavior of UHMWPE/alumina-zirconia composites for biomedical applications. *Eur. Polym. J.* **2019**, *120*, 109274. [\[CrossRef\]](#)
20. Yan, C.; Zhang, P.; Qin, Q.; Jiang, K.; Luo, Y.; Xiang, C.; He, J.; Chen, L.; Jiang, D.; Cui, W. 3D-printed bone regeneration scaffolds modulate bone metabolic homeostasis through vascularization for osteoporotic bone defects. *Biomaterials* **2024**, *311*, 122699. [\[CrossRef\]](#)
21. Antunovic, F.; Tolosa, F.; Klein, C.; Ocaranza, R. Polycaprolactone-based scaffolds for guided tissue regeneration in periodontal therapy: A systematic review. *J. Appl. Biomater. Funct. Mater.* **2023**, *21*, 22808000231211416. [\[CrossRef\]](#)
22. Xu, X.; Zhou, Y.; Zheng, K.; Li, X.; Li, L.; Xu, Y. 3D polycaprolactone/gelatin-oriented electrospun scaffolds promote periodontal regeneration. *ACS Appl. Mater. Interfaces* **2022**, *14*, 46145–46160. [\[CrossRef\]](#)
23. Khor, H.L.; Ng, K.W.; Schantz, J.-T.; Phan, T.-T.; Lim, T.C.; Teoh, S.-H.; Hutmacher, D. Poly (ϵ -caprolactone) films as a potential substrate for tissue engineering an epidermal equivalent. *Mater. Sci. Eng. C* **2002**, *20*, 71–75. [\[CrossRef\]](#)
24. Chunyan, Z.; Lan, C.; Jiajia, L.; Dongwei, S.; Jun, Z.; Huinan, L. In vitro evaluation of degradation, cytocompatibility and antibacterial property of polycaprolactone/hydroxyapatite composite coating on bioresorbable magnesium alloy. *J. Magnes. Alloys* **2022**, *10*, 2252–2265. [\[CrossRef\]](#)
25. Di Maro, M.; Pedraza, R.; Mosca Balma, A.; Gomez d’Ayala, G.; Poggetto, G.D.; Malucelli, G.; Roato, I.; Duraccio, D.; Mussano, F.; Faga, M.G. Influence of Dry-Mixing and Solvent Casting Blending Techniques on the Mechanical and Biological Behavior of Novel Biocompatible Poly (ϵ -caprolactone)/Alumina-Toughened Zirconia Scaffolds Obtained by 3D Printing. *J. Compos. Sci.* **2024**, *8*, 194. [\[CrossRef\]](#)
26. Waldner, C.; Hirn, U. Modeling liquid penetration into porous materials based on substrate and liquid surface energies. *J. Colloid Interface Sci.* **2023**, *640*, 445–455. [\[CrossRef\]](#)
27. Annamalai, M.; Gopinadhan, K.; Han, S.A.; Saha, S.; Park, H.J.; Cho, E.B.; Kumar, B.; Patra, A.; Kim, S.-W.; Venkatesan, T. Surface energy and wettability of van der Waals structures. *Nanoscale* **2016**, *8*, 5764–5770. [\[CrossRef\]](#)
28. Roato, I.; Genova, T.; Duraccio, D.; Ruffinatti, F.A.; Zanin Venturini, D.; Di Maro, M.; Mosca Balma, A.; Pedraza, R.; Petrillo, S.; Chinigò, G. mechanical and biological characterization of PMMA/Al₂O₃ composites for dental implant abutments. *Polymers* **2023**, *15*, 3186. [\[CrossRef\]](#)
29. Berta, G.N.; Di Scipio, F.; Yang, Z.; Oberto, A.; Abbadessa, G.; Romano, F.; Carere, M.E.; Ceccarelli, A.; Hirsch, E.; Mognetti, B. Chemical Oral Cancerogenesis Is Impaired in PI3K γ Knockout and Kinase-Dead Mice. *Cancers* **2021**, *13*, 4211. [\[CrossRef\]](#)
30. Stringer, C.; Wang, T.; Michaelos, M.; Pachitariu, M. Cellpose: A generalist algorithm for cellular segmentation. *Nat. Methods* **2021**, *18*, 100–106. [\[CrossRef\]](#)
31. Stringer, C.; Pachitariu, M. Cellpose3: One-click image restoration for improved cellular segmentation. *bioRxiv* **2024**. bioRxiv:2010.579780.
32. JCGM 100:2008; Evaluation of Measurement Data—Guide to the Expression of Uncertainty in Measurement, JCGM 100: 2008 GUM 1995 with Minor Corrections. Joint Committee for Guides in Metrology: Sèvres, France, 2008; p. 98.

33. Dietrich, C.F. *Uncertainty, Calibration and Probability: The Statistics of Scientific and Industrial Measurement*; Routledge: Abingdon, UK, 2017.
34. Babich, H.; Sedletcaia, A.; Kenigsberg, B. In vitro cytotoxicity of protocatechuic acid to cultured human cells from oral tissue: Involvement in oxidative stress. *Pharmacol. Toxicol.* **2002**, *91*, 245–253. [\[CrossRef\]](#)
35. Wong, P.-C.; Song, S.-M.; Tsai, P.-H.; Nien, Y.-Y.; Jang, J.S.-C.; Cheng, C.-K.; Chen, C.-H. Relationship between the surface roughness of biodegradable Mg-based bulk metallic glass and the osteogenetic ability of MG63 osteoblast-like cells. *Materials* **2020**, *13*, 1188. [\[CrossRef\]](#)
36. De Avila, E.D.; De Molon, R.S.; Palomari Spolidorio, D.M.; de Assis Mollo Jr, F. Implications of surface and bulk properties of abutment implants and their degradation in the health of periodontal tissue. *Materials* **2013**, *6*, 5951–5966. [\[CrossRef\]](#)
37. Bacakova, L.; Filova, E.; Parizek, M.; Ruml, T.; Svorcik, V. Modulation of cell adhesion, proliferation and differentiation on materials designed for body implants. *Biotechnol. Adv.* **2011**, *29*, 739–767. [\[CrossRef\]](#)
38. Poddar, D. Effect of Bulk and Surface Properties of Polymeric Scaffolds on Its Applications in Tissue Engineering. In *Tailored Functional Materials for Clean and Sustainable Development*; Apple Academic Press: Palm Bay, FL, USA, 2024; pp. 59–85.
39. Greiner, R.; Schwarzl, F. Thermal contraction and volume relaxation of amorphous polymers. *Rheol. Acta* **1984**, *23*, 378–395. [\[CrossRef\]](#)
40. Turner, B.N.; Strong, R.; Gold, S.A. A review of melt extrusion additive manufacturing processes: I. Process design and modeling. *Rapid Prototyp. J.* **2014**, *20*, 192–204. [\[CrossRef\]](#)
41. Bellini, A. *Fused Deposition of Ceramics: A Comprehensive Experimental, Analytical and Computational Study of Material Behavior, Fabrication Process and Equipment Design*; Drexel University: Philadelphia, PA, USA, 2002.
42. Shofner, M.L.; Lozano, K.; Rodríguez-Macías, F.J.; Barrera, E.V. Nanofiber-reinforced polymers prepared by fused deposition modeling. *J. Appl. Polym. Sci.* **2003**, *89*, 3081–3090. [\[CrossRef\]](#)
43. Puszka, A.; Sikora, J.W. New Segmented Poly (Thiourethane-Urethane) s Based on Poly (ϵ -Caprolactone) Diol Soft Segment: Synthesis and Characterization. *Materials* **2022**, *15*, 4940. [\[CrossRef\]](#)
44. Carvalho, A.; Grenho, L.; Fernandes, M.H.; Daskalova, A.; Trifonov, A.; Buchvarov, I.; Monteiro, F.J. Femtosecond laser microstructuring of alumina toughened zirconia for surface functionalization of dental implants. *Ceram. Int.* **2020**, *46*, 1383–1389.
45. Yang, Y.; Knust, S.; Schwiderek, S.; Qin, Q.; Yun, Q.; Grundmeier, G.; Keller, A. Protein adsorption at nanorough titanium oxide surfaces: The importance of surface statistical parameters beyond surface roughness. *Nanomaterials* **2021**, *11*, 357. [\[CrossRef\]](#)
46. Martinez, M.A.F.; Balderrama, Í.d.F.; Karam, P.S.B.H.; de Oliveira, R.C.; de Oliveira, F.A.; Grandini, C.R.; Vicente, F.B.; Stavropoulos, A.; Zangrando, M.S.R.; Sant’Ana, A.C.P. Surface roughness of titanium disks influences the adhesion, proliferation and differentiation of osteogenic properties derived from human. *Int. J. Implant Dent.* **2020**, *6*, 46. [\[CrossRef\]](#) [\[PubMed\]](#)
47. Olivares-Navarrete, R.; Hyzy, S.L.; Berg, M.E.; Schneider, J.M.; Hotchkiss, K.; Schwartz, Z.; Boyan, B.D. Osteoblast lineage cells can discriminate microscale topographic features on titanium–aluminum–vanadium surfaces. *Ann. Biomed. Eng.* **2014**, *42*, 2551–2561. [\[CrossRef\]](#)
48. Bellon, B.; Pippenger, B.; Stähli, A.; Degen, M.; Parisi, L. Cementum and enamel surface mimicry influences soft tissue cell behavior. *J. Periodontal Res.* **2024**, 1–13. [\[CrossRef\]](#) [\[PubMed\]](#)
49. Saberian, E.; Jenča, A.; Seyfaddini, R.; Jenča, A.; Zare-Zardini, H.; Petrášová, A.; Jenčová, J. Comparative Analysis of Osteoblastic Responses to Titanium and Alumina-Toughened Zirconia Implants: An In Vitro Study. *Biomolecules* **2024**, *14*, 719. [\[CrossRef\]](#)
50. Nazarov, D.; Ezhov, I.; Yudintceva, N.; Shevtsov, M.; Rudakova, A.; Kalganov, V.; Tolmachev, V.; Zharova, Y.; Lutakov, O.; Kraeva, L. Antibacterial and osteogenic properties of Ag nanoparticles and Ag/TiO₂ nanostructures prepared by atomic layer deposition. *J. Funct. Biomater.* **2022**, *13*, 62. [\[CrossRef\]](#)
51. Riivari, S.; Närvä, E.; Kangasniemi, I.; Willberg, J.; Närhi, T. Epithelial cell attachment and adhesion protein expression on novel in sol TiO₂ coated zirconia and titanium alloy surfaces. *J. Biomed. Mater. Res. Part B: Appl. Biomater.* **2022**, *110*, 2533–2541. [\[CrossRef\]](#)

Disclaimer/Publisher’s Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.