



ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

Realization of the first sub-shot-noise wide field microscope

Original

Realization of the first sub-shot-noise wide field microscope / Samantaray, NIGAM LAHARI; Ruo-Berchera, Ivano; Meda, Alice; Genovese, Marco. - In: LIGHT, SCIENCE & APPLICATIONS. - ISSN 2047-7538. - 6:7(2017), p. e17005. [10.1038/lisa.2017.5]

Availability:

This version is available at: 11696/56865 since: 2021-02-01T13:09:33Z

Publisher:

Springer Nature

Published

DOI:10.1038/lisa.2017.5

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

ORIGINAL ARTICLE

Realization of the first sub-shot-noise wide field microscope

Nigam Samantaray^{1,2}, Ivano Ruo-Berchera¹, Alice Meda¹ and Marco Genovese^{1,3}

Recently, several proof of principle experiments have demonstrated the advantages of quantum technologies over classical schemes. The present challenge is to surpass the limits of proof of principle demonstrations to approach real applications. This letter presents such an achievement in the field of quantum enhanced imaging. In particular, we describe the realization of a sub-shot-noise wide field microscope based on spatially multi-mode non-classical photon number correlations in twin beams. The microscope produces realtime images of 8000 pixels at full resolution, for a 500 μm^2 field of view, with noise reduced to 80% of the shot noise level (for each pixel), which is suitable for absorption imaging of complex structures. By fast post-elaboration, specifically applying a quantum enhanced median filter, the noise can be further reduced (to <30% of the shot noise level) by setting a trade-off with the resolution, thus achieving the best sensitivity per incident photon reported in absorption microscopy.

Light: Science & Applications (2017) 6, e17005; doi:10.1038/lsa.2017.5; published online 14 July 2017

Keywords: imaging; microscopy; parametric down conversion; sub-shot-noise

INTRODUCTION

Sensitivity in standard optical imaging and sensing, such as those exploiting classical illuminating fields, is fundamentally lower bounded by the shot noise, which is the inverse square root of the number of photons used. Surpassing such a limit is particularly effective when there is a constraint on the usable optical power, for example, a limit determined by the damage threshold of the sample¹, the stress of the optical elements², or alteration of chemical and biological photo sensitive process and, most fundamentally, quantum back-action³. Following the seminal works of Caves⁴, which have demonstrated how squeezed light can improve the sensitivity in interferometry, non-classical states of light have long been considered to overcome shot noise, thus prompting deep theoretical investigations and many proposed schemes^{5–9}.

The experimental possibility of generating two-photon entangled states¹⁰ (such as NOON states with $N=2$) and the availability of single-photon detectors have enabled the demonstration of the potential of quantum enhanced sensing, aimed at reaching the fundamental Heisenberg limit in phase-contrast polarization microscopy^{11,12}, magnetic field sensing¹³ and solution concentration measurement¹⁴.

However, almost all previous experimental results have been proof of principle demonstrations⁶, because of the difficulties in generating high photon number entangled states and a high photon flux (comparable to that used in classical schemes) and protecting them from decoherence up to the detection, thus limiting the optical and detection losses¹⁵.

Some remarkable results have been obtained because of recent progress in the generation of a highly non-classical single mode¹⁶ and

few modes squeezed states¹⁷; these have been successfully implemented in gravitational wave interferometry², for particle tracking in a biological environment^{18,19}, and to some extent, for beam displacement measurement^{20,21} and optical magnetometry²².

Most quantum-enhanced imaging and sensing protocols have been obtained exploiting single or few spatial modes of the quantum probe beam, such as in the case of squeezing, and with single photon detection in schemes based on two-photon correlated states. In both cases, only one parameter of the system, namely a single point of the sample, can be probed in a single run. The reconstruction of the sample as a whole requires time consuming scanning and accumulation of many detection windows. Instead, the exploitation of a high number, namely thousands, of modes in the same run is the requirement for quantum enhanced wide field imaging. In practice, the number of spatial details of a structure that can be probed at the same time is determined by the number of spatial modes enclosed in the illuminating field. Indeed, one of the challenges in quantum optics and quantum enhanced imaging is to generate and efficiently detect highly non-classical features in a multi-mode regime^{23–28}. The first proof of principle of a quantum enhanced imaging protocol exploiting this parallelism was reported in Ref. 29, following the proposal of Ref. 30, even if the average enhancement and the poor spatial resolution were not sufficient for any practical purposes, such as in absorption microscopy, in which the technique is naturally addressed. In this paper, we address this point by reporting the realization of a sub-shot-noise (SSN) microscope exploiting thousands of spatial modes, detected independently by the same number of pixels of a charge-coupled-device (CCD) camera operated in the linear (non-amplified)

¹INRIM, Strada delle Cacce 91, I-10135 Torino, Italy; ²Politecnico di Torino, Corso Duca degli Abruzzi, 24, I-10129 Torino, Italy and ³INFN, Via P. Giuria 1, I-10125 Torino, Italy
Correspondence: A Meda, Email: a.meda@inrim.it

Received 29 August 2016; revised 16 January 2017; accepted 24 January 2017; accepted article preview online 28 January 2017

regime. Thousands of photons per pixel are detected in the exposure time of a single shot. Therefore, the microscope operates in a wide-field regime (no scanning is required) and is suitable for dynamic imaging. The microscope is based on the non-classical and spatially multimode correlations of squeezed vacuum that is naturally generated by a traveling wave parametric amplifier in both low and high gain regimes^{24,26,30–32}. The noise of the image, formed by the probe beam interacting with the sample, is locally reduced by pixel-by-pixel subtraction of the correlated noise pattern measured on the other beam³⁰. Moreover, we introduce the concept of a quantum enhanced median filter³³: quantum noise reduction at different spatial scales can be naturally combined with the statistical noise smoothing used in every standard image processing algorithm, thus resulting in an impressive overall enhancement in object recognition.

We attain a significant improvement in sensitivity with respect to that of any classical absorption microscopy system at the same illumination level. Our present results outperform the previous proof of principle demonstrations^{29,34}, substantially improving the resolution by a factor of 10–100 (depending on the sensitivity level), in terms of both pixel count and the size of the imaged details of the sample.

Wide field microscopy is the simplest, fastest, least expensive and oldest imaging modality used, for example, for live cell imaging. It has the advantage of requiring the lowest photon dose, especially for absorbed light imaging. The lowest photon dose that achieves a measurable metric for the experimental question should be used, for instance to ensure that the cellular processes under investigation are not shifted to an alternate pathway due to environmental stress³⁵. Indeed, the results presented here could potentially have immediate applications in many fields, such as biology and biochemistry. Furthermore, the comparison with the reference beam can be used to provide the absolute value of the absorption, thus providing the possibility for quantitative analysis of the related properties. With small modifications (essentially in the data processing), our technique can also serve as a basis for enhancing sensitivity in schemes with different goals: for example, ghost imaging^{36–41}, detection and imaging in the presence of environmental or electronic noise background^{42,43} and accurate characterization of retina rod-cell response to single photon stimulation⁴⁴. Finally, this new capability can also be used in quantum radiometry, for example, for the absolute calibration of detectors with spatial resolution, as demonstrated in Refs 45–48.

MATERIALS AND METHODS

In absorption wide-field imaging, as in standard microscopy, a probe illuminates the sample all at once and the transmitted pattern is imaged by the detector, typically as the pixel array of a camera. The intensity measured by each pixel, N_α , here expressed in number of photons, has a mean expectation value $\langle N_\alpha \rangle = (1 - \alpha)\langle N \rangle$, where $\langle N \rangle$ is the mean number of detected photons per pixel in absence of the sample and α is the absorption coefficient. The photon noise of the measurement can be obtained by modeling the absorption as the action of a beam splitter of transmittance $1 - \alpha$ on the beam with initial variance $\langle \Delta^2 N \rangle$ (Ref. 49), thus resulting in the form $\langle \Delta^2 N_\alpha \rangle = [(1 - \alpha)^2(F - 1) + 1 - \alpha]\langle N \rangle$, where $F = \langle \Delta^2 N \rangle / \langle N \rangle$ is the Fano factor in absence of the sample. The value $F = 1$ establishes a bound between classical and non-classical photon statistics. In particular, F is lower bounded by the unity for classical states, whereas specific non-classical states may have sub-Poissonian photon statistics, that is, $0 \leq F < 1$. The uncertainty of the absorption estimation in the

direct (DR) imaging scheme is therefore

$$\Delta\alpha_{\text{DR}} = \frac{\sqrt{\langle \Delta^2 N_\alpha \rangle}}{|\partial_\alpha \langle N_\alpha \rangle|} = \sqrt{\frac{(1 - \alpha)^2(F - 1) + 1 - \alpha}{\langle N \rangle}} \quad (1)$$

The limit of the sensitivity for a classical probe ($F = 1$) is $\Delta\alpha = \sqrt{(1 - \alpha)/\langle N \rangle}$, representing, for small absorption, the shot-noise limit with the typical scaling of the inverse square root of the number of photons. However, by inspecting Equation (1), it is clear that non-classical optical fields with Fano factor smaller than one allow beating the shot noise limit. We note that the Fano factor, appearing in Equation (1), is usually deteriorated with respect to the value of F_0 for the unperturbed (pure) state, due to optical losses, including detector quantum efficiency. In particular, one obtains $F = \eta F_0 + 1 - \eta$ where we define the overall detection probability $0 \leq \eta \leq 1$. Thus, the non-classical behavior, in terms of noise reduction, has a lower bound of $F_{\text{loss}} = 1 - \eta$. It must be emphasized that splitting a single mode beam into n pixels leads to a detection probability of the order of $\eta \leq 1/n$ for each pixel, ruling out the possibility of using single modes for SSN wide-field imaging for any reasonable number of pixels. Thus, the necessity of having many non-classical spatial modes is evident, each addressing a single pixel with limited losses. Even if sub-Poissonian light beams were obtained as single or few modes, it is not completely obvious how to generate a beam with a high number of sub-Poissonian modes and simultaneously detect them. On the other side, it is relatively simple to produce a pair of beams that are (individually) spatially incoherent but locally correlated at the quantum level, by means of traveling wave parametric amplifier in the spontaneous regime. Even if the fluctuations of a single spatial mode in one beam are super-Poissonian, these fluctuations are perfectly replicated in the correlated mode of the second beam because of photon number entanglement. This correlation is verified for all of the wide range of localized transverse spatial modes. The degree of correlation and its non-classical features can be quantified on the basis of the noise reduction factor (NRF) $\sigma = \langle \Delta^2(N_1 - N_2) \rangle / \langle N_1 + N_2 \rangle$ (Refs 23,24,26,30–32,49–52), measured for a pair of pixels by collecting correlated spatial modes. The NRF indeed represents the equivalent of the Fano factor for bipartite state, in which the shot noise level becomes $\langle N_1 + N_2 \rangle = 2\langle N \rangle$. While for classical beams the NRF is lower bounded by 1, quantum correlation can lead to $0 \leq \sigma < 1$. In particular, in the presence of losses, $\sigma = 1 - \eta$.

Whereas twin beams in a single spatial mode have been used to demonstrate SSN absorption measurements in a double-beam scheme^{53,54}, parametric down-conversion (PDC) multi-mode quantum correlations can be used for wide-field SSN imaging³⁰.

The object is placed in one beam, and the second beam is used as a reference. Note that the double-beam (or double-path) approach is commonly used in imaging and spectroscopy involving faint absorptions, because it enables the canceling out of classical (super-Poissonian) noise and provides a direct estimation of the absolute transmittance (absorption) by instantaneous comparison with the unperturbed reference beam. We consider the intensity difference between two correlated pixels of the two beams, whose expectation value is $\langle N_1 - N_{2,\alpha} \rangle = \alpha\langle N \rangle$ (we assume balanced beams, $\langle N_1 \rangle = \langle N_2 \rangle = \langle N \rangle$). In this context, the noise can be expressed in terms of the noise reduction factor in absence of the sample, in the form $\langle \Delta^2(N_1 - N_{2,\alpha}) \rangle = [\alpha^2(F - 1) + \alpha + 2\sigma(1 - \alpha)]\langle N \rangle$ (see Ref. 30).




Figure 1 Experimental setup. (a) Schematic of the experiment. (b) Image of the single shot acquired by the CCD camera in 100 ms. The pixel size is $39 \mu\text{m}$, corresponding to $5\text{-}\mu\text{m}$ resolution in the focal (object) plane. (c) 2D map of the NRF for $5\text{-}\mu\text{m}$ resolution in the focal (object) plane.

Therefore, the absorption uncertainty in the differential (DF), both classical and quantum, imaging scheme is

$$\Delta\alpha_{\text{DF}} = \frac{\sqrt{\langle \Delta^2(N_1 - N_{2,\alpha}) \rangle}}{|\partial_\alpha \langle N_{-\alpha} \rangle|} = \sqrt{\frac{\alpha^2(F-1) + \alpha + 2\sigma(1-\alpha)}{\langle N \rangle}} \quad (2)$$

The performance of the classical differential (DC) imaging scheme is derived from Equation (2) by substituting $\sigma=1$. In the situation of interest, the absorption is so small that the term $\alpha^2(F-1)$ is negligible even in presence of classical super-Poissonian noise of the source ($F > 1$), and the uncertainty in the differential classical imaging scheme becomes $\Delta\alpha_{\text{DC}} = \sqrt{(2-\alpha)/\langle N \rangle}$, which is only a factor of $\sqrt{2}$ larger than shot-noise-limited direct imaging. Under the same condition, the quantum enhancement provided by SSN correlations with $\sigma < 1$ is also quantified in terms of signal-to-noise ratio, $\text{SNR} = \alpha/\Delta\alpha$, by using Equations (1) and (2), as

$$\frac{\Delta\alpha_{\text{SSN}}}{\Delta\alpha_{\text{DC}}} = \frac{\text{SNR}_{\text{DC}}}{\text{SNR}_{\text{SSN}}} = \sqrt{\frac{\alpha + 2\sigma(1-\alpha)}{2-\alpha}} \approx \sqrt{\sigma} \quad (3a)$$

$$\frac{\Delta\alpha_{\text{SSN}}}{\Delta\alpha_{\text{DR}}} = \frac{\text{SNR}_{\text{DR}}}{\text{SNR}_{\text{SSN}}} = \sqrt{\frac{\alpha + 2\sigma(1-\alpha)}{1-\alpha}} \approx \sqrt{2\sigma} \quad (3b)$$

The advantage with respect to the differential classical scheme appears when $\sigma < 1$, whereas a more strict condition, $\sigma < 1/2$, is needed to beat the direct (shot-noise-limited) scheme.

The setup for the SSN microscope scheme is sketched in Figure 1a. A CW laser beam (100 mW at $\lambda_p = 405 \text{ nm}$) pumps a 1-cm type-II beta-barium-borate nonlinear crystal, in which two correlated beams are generated. The far field of the emission, where spatial correlation

occurs, is realized at the focal plane of a lens with $f_{\text{FF}} = 1 \text{ cm}$ focal length. Then, the far field plane is imaged (magnification factor $M = 7.8$) to the detection plane by means of a second lens system with $f_{\text{IM}} = 1.6 \text{ cm}$. The detector is a CCD camera Princeton Inst. Pixis 400BR Excelon, operating in linear mode (no electro-multiplication gain), with high quantum efficiency ($> 95\%$ at 810 nm), 100% fill factor and low noise [read noise is few $e^-/(\text{pix} \cdot \text{frame})$]. The size of the physical pixels of the camera is $13 \mu\text{m}$, but here we group them by a 3×3 hardware binning. Hereafter, if not explicitly indicated, the single elementary pixel is intended to have a linear size of $39 \mu\text{m}$, thus causing a sufficient reduction of the effect of the read noise for our purposes. A spectral selection is performed with two identical interferential filters ($800 \pm 20 \text{ nm}$, with transmission of 99%), one just after the crystal and the other mounted on the camera. A test sample, with absorption $\alpha = 1\%$ representing the Greek letter Φ (size $300 \times 400 \mu\text{m}$) is realized by a few nanometers thick titanium deposition on a coated glass slide.

When inserted, the letter intercepts one beam at the focal plane of the far field lens whereas the slide extends to the second beam (in the far field, the centers of the two beams are separated by 1.0 mm). It is important to note that the acquisition time of a single shot is typically 10^{11} times the coherence time of the PDC process (on the order of 10^{-12} s). Since the number of photons per pixel is $\sim 10^3$, the occupation of the single temporal mode is 10^{-8} photons/mode. Therefore, the statistics of the noise should be Poissonian, that is, shot-noise limited.

In the crystal, a photon of the pump beam is converted into a pair of photons with lower frequency, thus fulfilling the energy and momentum conservation (phase matching-condition). In particular,

if the pump is approximated by a plane wave with transverse momentum $\mathbf{q}=0$, the two photons of a pair must fulfill $\mathbf{q}_1+\mathbf{q}_2=0$, thus being emitted with opposite transverse momenta. The lens maps momentum \mathbf{q} at wavelength λ into point $\mathbf{x}=\lambda f\mathbf{q}/2\pi$ at the focal plane, thus leading degenerate photons (having $\lambda_1=\lambda_2=\lambda_d=2\lambda_p=810$ nm) being found in symmetric positions $\mathbf{x}_1=-\mathbf{x}_2$ with respect to the pump direction, as shown in Figure 1a. Two symmetric pixels of the camera should always detect the same number of photons. Moreover, even if the emission is broadband both in frequency and momentum, phase matching conditions establish a relationship between the wavelength and the photon direction, specifically the modulus of the transverse momentum, so that photons are emitted in concentric cones, each corresponding to a certain wavelength. In Type-II PDC, the correlated photons have orthogonal polarizations, and the emission cones have different centers. The intensity distribution, detected in a 100-ms shot, is shown in Figure 1b. The dotted circumferences approximately represent the wavelength distribution in the two orthogonally polarized beams. By selecting two symmetric regions, *A* and *B*, around degeneracy (represented by red squares in the picture), one expects each pixel of *A* to be non-classically correlated with a corresponding symmetric pixel in *B*. In practice, there are two important limiting factors to this imaging system that must be considered. One limits the field of view, and the other lower bounds the spatial scale at which the noise subtraction can be efficiently performed:

- The trade-off between field of view and NRF: Near the degeneracy wavelength for correlated photons, $\lambda_1=\lambda_d+\Delta\lambda$ and $\lambda_2=\lambda_d-\Delta\lambda$; the point-to-point correlation in the far field becomes $\mathbf{x}_1+\mathbf{x}_2\cong\frac{2\Delta\lambda}{\lambda_d}\mathbf{x}_1$. Therefore, the center of symmetry of the correlations is $\mathbf{x}=0$ only at the degenerate wavelength. As long as one moves from the degeneracy, the center of symmetry shifts proportionally to $\frac{\Delta\lambda}{\lambda_d}$. After the pixel grid has been positioned to be symmetric with respect to $\mathbf{x}=0$, only a relatively small spectral bandwidth around degeneracy can be tolerated (in our case, ~ 40 nm). This reflects on the available angular bandwidth, see Figure 1b, which corresponds to an approximately $500\times 500\ \mu\text{m}$ field of view in the focal plane. Figure 1c shows a map of the NRF obtained by subtracting locally, pixel-by-pixel, the two regions *A* and *B*. The best NRF occurs along the degenerate ring, and far from it, the correlation slightly decreases because the corresponding pixels of the two regions no longer perfectly intercept correlated directions. Moreover, right-hand part shows a further increase of the NRF, which our experience suggests can be ascribed to a small aberration in the optical system, and which we are able to decrease but not completely suppress. This technical issue can be solved by a careful analysis and realization of an ad-hoc optical system.
- The trade-off between spatial resolution and NRF: In practice, rather than a plane wave, the pump is a Gaussian beam propagating along the *z* direction, with a waist w_p , and thus with a transverse momentum distribution centered at $\mathbf{q}=0$ with bandwidth $\Delta\mathbf{q}\sim 2/w_p$. Transverse momentum conservation $\mathbf{q}_1+\mathbf{q}_2=0\pm\Delta\mathbf{q}$ leads to a less strict position correlation in the far field, $\mathbf{x}_1+\mathbf{x}_2=\pm\Delta\mathbf{x}$, where the relative uncertainty of the photon position is $\Delta\mathbf{x}\cong 2r\sim\lambda f/\pi w_p$. A measurement of this spatial uncertainty is provided in Figure 2, representing the spatial cross correlation function of the noise patterns of two symmetric regions of 40×40 physical pixels. The cross section is Gaussian with two slightly different FWHM in the horizontal and vertical axes, estimated to be $2r_y=5.6(0.2)\ \mu\text{m}$ and $2r_x=6.8(0.2)\ \mu\text{m}$, respectively. It is clear that two symmetric pixels detect most of the correlated photons only if their size *L* is larger than this uncertainty (see Ref. 32).




Figure 2 Spatial cross correlation function. The 2D map represents the value of the correlation coefficient between two regions of 40×40 pixels, chosen approximately in symmetric positions at the detection plane. Here, we exploited the full resolution of the camera (physical pixel of $13\ \mu\text{m}$), which corresponds to $1.7\ \mu\text{m}$ in the object plane. The peak represents the position in which the regions are well correlated, pixel-by-pixel. Shifting one of the regions in the pixel grid more than the spatial jitter of the correlated photon makes the correlation coefficient drop to zero. The vertical and horizontal sections are shown in the left-hand-side and bottom graphs, respectively, with their fitting Gaussian functions and the indication of the FWHMs. Corr. coeff, correlation coefficient.

RESULTS AND DISCUSSION

The NRF can be described by the function $\sigma=1-\eta_0\cdot\eta_{\text{coll}}$, in which the detection probability η is split into the term η_0 , representing the transmission-detection efficiency of the optical path, and $0<\eta_{\text{coll}}<1$, representing the collection efficiency of correlated photons. The efficiency η_{coll} is a monotonically increasing function of the ratio $L/2r$, which reaches the asymptotic value $\eta_{\text{coll}}\sim 1$ for $L\gg 2r$ (Supplementary Information). Thus, at different resolution scales, given by the size of pixels or, more generally, by the spatial scale in which the signal is integrated, the noise reduction factor and thus the SNR, according to Equation (3a) and (3b), are different. The spatial information at different scales can be recovered starting from a high resolution image in which the pixels size x_{pix} can be smaller than the correlation area and by averaging the signal in groups of $d\times d$ pixels so that $L=d\cdot x_{\text{pix}}$. This is particularly useful, given that it can be interpreted as a quantum-enhanced version of the standard median filter used for noise correction in classical imaging. In the classical median filter, the value of a pixel is replaced with the mean value of its $d\times d$ neighborhood, thus enabling a reduction of the noise simply by a statistical cancellation. Of course, this comes at the expense of filter details smoothing. If the median filter is applied to the noise-subtracted image, the two effects, namely the statistical smoothing and the photon noise subtraction, combine, thereby allowing a net improvement of faint object recognition. Figure 3 shows an example of the application of the median filter to a direct image, a classical differential image, and a SSN image. The upper-right panel represents the shape of the object obtained by averaging over 300 shots at full resolution ($L=5\ \mu\text{m}$). The image represents the absorption coefficient, assuming values around zero outside the Φ and value around $\alpha=0.01$ inside the Φ . The other panels present single shot images at different resolution scales. The SSN images are obtained by the subtraction of




Figure 3 Application of a median filter to a single shot image for different integration scales d . The direct (DR) image, the differential classical (DC) image, and the SSN image are compared in each panel for the same value of d . Upper-left panel is the image of the object after averaging over 300 shots.

the correlated noise from the direct image. The DC images are simulated by subtracting an uncorrelated (but shot-noise-limited) noise pattern, for example, a noise region shifted more than $2r$ from the correlated position. Notably, the single shot image of the sample is completely hidden by the shot noise at the full resolution, $d=1$. When the median filters are applied, $d>1$, the shape of the sample clearly emerges, especially for the SSN case. As discussed before, the advantage of the SSN image increases with the scale d of the median filter. The DC image is the poorest because it contains twice the shot noise. Notably, DC imaging is advantageous compared with the DR imaging if classical super-Poissonian noise is present, in particular when $F>2$.

The exposure time of the single shot in our setup is 100 ms, and the read-out time of the pixel matrix is few hundreds of milliseconds. In a few seconds, it is possible to realize a SSN movie (See Supplementary Movie) in which the sample is simply translated by a micrometer stage during the acquisition. Even in this case, the edges of the moving object are easier to follow in the SSN imaging box (right-hand side window) than in the DC imaging (central window) or in the DR imaging box (left-hand side window), thus demonstrating that our technique is suitable for dynamic imaging. With a more powerful or pulsed pump laser and the faster operating modes achievable by commercial cameras, rate of hundreds of frame/s should be reasonably attained.

The statistical analysis of the SNR is performed by acquiring $\mathcal{N} = 300$ shots with and without the sample. The shots without the sample are used to estimate the NRF and Fano factor at different scales and to eliminate static intensity and efficiency gradients by a standard flat field algorithm. Each shot with the sample contains a DR shot-noise limited image and the correlated noise. The NRF of the n th shot is evaluated experimentally through spatial statistics over the ensemble of the correlated pixel pairs belonging to regions A and B , namely $\sigma(n) = \frac{V_x[N_A^{(n)}(\mathbf{x}) - N_B^{(n)}(-\mathbf{x})]}{E_x[N_A^{(n)} + N_B^{(n)}]}$, where $E_x[N(\mathbf{x})] = (1/\mathcal{M})\sum_x N(\mathbf{x})$ is the mean value of the \mathcal{M} pixel of the region and

$V_x[N(\mathbf{x})] = E_x[N^2(\mathbf{x})] - E_x[N(\mathbf{x})]^2$ is the variance. The average NRF over the 300 values is reported in the graph in Figure 4 for different resolution scales in the object plane (see also Supplementary Information). In particular, we note that already for a resolution of $5 \mu\text{m}$, comparable with the correlation FWHM of the spatial correlation function, the system reaches an NRF of $\sigma=0.8$, thus providing better performance than DC imaging (see Equation (3a)) in a matrix of >8000 pixels. For a resolution of $15 \mu\text{m}$ the NRF is below 0.5, which is the threshold to surpass the performance of the DR, SSN limited, imaging (see Equation (3b)). Similarly, the Fano factor is evaluated and reported in Figure 4, thus demonstrating the Poissonian character of the light statistics.

The SNR is estimated over a stripe of pixels of the images after the application of the median filter (in particular, we consider a vertical stripe inside the main axis of the letter Φ). We label $\alpha^{(n)}(\mathbf{x})$ as the absorption value of the pixel in position \mathbf{x} of the n th shot. First, the SNR for each position is evaluated as $\text{SNR}(\mathbf{x}) = E_n[\alpha(\mathbf{x})]/V_n[\alpha(\mathbf{x})]^{1/2}$, where $E_n[\alpha(\mathbf{x})] = (1/\mathcal{N})\sum_{n=1}^{\mathcal{N}} \alpha^{(n)}(\mathbf{x})$ is the experimental temporal average of the absorption and $V_n[\alpha(\mathbf{x})]$ is the variance. Then, the spatial average of the SNR(\mathbf{x}) for \mathbf{x} belonging to the vertical stripe is evaluated. The experimental results, demonstrating the advantages of the quantum noise subtraction, are reported in Figure 4. The data are compared with the theoretical predictions obtained by substituting the estimated NRF in the theoretical expression of Equation (3a) and (3b). The SNR improvement of the SSN imaging is slightly higher than expected, with respect to both DR and DC imaging, because the SNR is evaluated on a vertical stripe close to the degenerate wavelength region, where the NRF is slightly lower than the average (see Figure 1c and the related discussion).

CONCLUSIONS

In summary, we realized the first sub-shot-noise wide field microscope, demonstrating a noise reduction of 20% below the shot noise for each resolution cell (pixel) of $5 \mu\text{m}$ in a matrix of ~ 8000 pixels. This value is sufficient for wide-field imaging of




Figure 4 Experimental noise reduction factor (NRF) and signal-to-noise ratio (SNR) as a function of the resolution in the focal (object) plane L (or equivalently as a function of the binning d of the median filter, upper scale). Red dots represent the average of the NRF in a region of the same size as the Φ object, that is $400 \times 300 \mu\text{m}$. The black dots are the SNR of the SSN images normalized to that of the direct images. For $L \geq 15 \mu\text{m}$, the quantum protocol has advantages. Analogously, the blue series demonstrates that the advantage of the SSN imaging with respect to the differential classical imaging is present at any spatial resolution and reaches values of more than 80%. Solid lines correspond to the quantum enhancement predicted by Equation (3a) and (3b), when the estimated values of the NRF are considered.

complex structures. Increasing the spatial scale of the details, the noise reduction factor improves accordingly; for example, a resolution of $25 \mu\text{m}$ is 62% below shot noise and for $50 \mu\text{m}$ is 72% on average below shot noise level. At this last scale, for example, the SNR is almost twice that of the classical differential technique for the same illumination level, whereas the sensitivity improvement with respect to classical direct imaging (obtained by a single, shot noise limited beam) is $\sim 30\%$. Equivalently it allows one to maintain the same SNR reducing almost four times the illumination level compared with the DC imaging and almost twice compared with DR imaging. In general, to have the SNR of DC (or of DR) imaging, the photon number of the quantum beam can be decreased by a factor σ (2σ for DR). We showed that the decrease in the quantum noise at different scales is fully compatible with, and actually improves, the standard noise reduction techniques, on the basis of a posteriori elaborations of the image taken at full resolution. We demonstrated this important point with a sample made of a ultra-thin metallic depositions on a glass slide. These performance levels represent a breakthrough, filling the gap between the proof-of-principle of quantum enhanced scheme and a system suitable for applications^{6,55}.

We believe that our technique has the potential for wide-spread use in absorption microscopy. The spatial resolution can be further improved (nothing prevents reaching the Abbe limit), and the range of applicability can be extended by engineering highly non-classical bright squeezed vacuum sources in a pulsed regime⁵².

CONFLICT OF INTEREST

The authors declare no conflict of interest.

AUTHOR CONTRIBUTIONS

IRB, AM and MG conceived the idea of the experiment, which was discussed and designed with input from all authors. NS, IRB, AM realized the experimental setup and collected the data in INRIM quantum optics labs (coordinated by MG). All authors discussed the results and contributed to the writing of the paper.

ACKNOWLEDGEMENTS

This work was supported by the MIUR Project 'Premiale P5'. We thank Matteo Fretto for the micro-fabrication of the object used in the measurement, F Scafrimuto for the help in a preliminary characterization of the system, and IP Degiovanni for fruitful discussions and advises.

- 1 Taylor MA, Bowen WP. Quantum metrology and its application in biology. *Phys Rep* 2016; **615**: 1–59.
- 2 The LIGO Scientific Collaboration. A gravitational wave observatory operating beyond the quantum shot-noise limit. *Nat Phys* 2011; **7**: 962–965.
- 3 Giovannetti V, Lloyd S, Maccone L. Quantum-enhanced measurements: beating the standard quantum limit. *Science* 2004; **306**: 1330–1336.
- 4 Caves CM. Quantum-mechanical noise in an interferometer. *Phys Rev D* 1981; **23**: 1693–1708.
- 5 Demkowicz-Dobrzański R, Jarzyna M, Kołodyński J. Quantum limits in optical interferometry. *Prog Opt* 2015; **60**: 345–435.
- 6 Genovese M. Real applications of quantum imaging. *J Opt* 2016; **18**: 073002.
- 7 Kolobov MI. *Quantum Imaging*. New York: Springer; 2007.
- 8 Nabors CD, Shelby RM. Two-color squeezing and sub-shot-noise signal recovery in doubly resonant optical parametric oscillators. *Phys Rev A* 1990; **42**: 556–559.
- 9 Souto Ribeiro PH, Schwob C, Maître A, Fabre C. Sub-shot-noise high-sensitivity spectroscopy with optical parametric oscillator twin beams. *Opt Lett* 1997; **22**: 1893–1895.
- 10 Matthews CFJ, Alberto P, Bonneau D, O'Brien JL. Heralding two-photon and four-photon path entanglement on a chip. *Phys Rev Lett* 2011; **107**: 163602.
- 11 Ono T, Okamoto R, Takeuchi S. An entanglement-enhanced microscope. *Nat Commun* 2013; **4**: 2426.
- 12 Israel Y, Rosen S, Silberberg Y. Supersensitive polarization microscopy using NOON states of light. *Phys Rev Lett* 2014; **112**: 103604.
- 13 Wolgramm F, Vitelli C, Beduini FA, Godbout N, Mitchell MW. Entanglement-enhanced probing of a delicate material system. *Nat Photonics* 2013; **7**: 28–32.
- 14 Crespi A, Lobino M, Matthews JCF, Politi A, Neal CR *et al*. Measuring protein concentration with entangled photons. *Appl Phys Lett* 2012; **100**: 233704.
- 15 Demkowicz-Dobrzański R, Kołodyński J, Guță M. The elusive Heisenberg limit in quantum-enhanced metrology. *Nat Commun* 2012; **3**: 1063.
- 16 Mehmet M, Ast S, Eberle T, Steinlechner S, Vahlbruch H *et al*. Squeezed light at 1550 nm with a quantum noise reduction of 12.3 dB. *Opt Express* 2011; **19**: 25763–25772.
- 17 Armstrong S, Morizur JF, Janousek J, Hage B, Treps N *et al*. Programmable multimode quantum networks. *Nat Commun* 2012; **3**: 1026.
- 18 Taylor MA, Janousek J, Daria V, Knittel J, Hage B *et al*. Biological measurement beyond the quantum limit. *Nat Photonics* 2013; **7**: 229–233.
- 19 Taylor MA, Janousek J, Daria V, Knittel J, Hage B *et al*. Subdiffraction-limited quantum imaging within a living cell. *Phys Rev X* 2014; **4**: 011017.
- 20 Treps N, Grosse N, Bowen WP, Fabre C, Bachor HA *et al*. A quantum laser pointer. *Science* 2003; **301**: 940–943.
- 21 Pooser RC, Lawrie B. Ultrasensitive measurement of microcantilever displacement below the shot-noise limit. *Optica* 2015; **2**: 393–399.
- 22 Wolgramm F, Cerè A, Beduini FA, Predojević A, Koschorreck M *et al*. Squeezed-light optical magnetometry. *Phys Rev Lett* 2010; **105**: 053601.
- 23 Brida G, Caspani L, Gatti A, Genovese M, Meda A *et al*. Measurement of sub-shot-noise spatial correlations without background subtraction. *Phys Rev Lett* 2009; **102**: 213602.
- 24 Jedrkiewicz O, Jiang YK, Brambilla E, Gatti A, Bache M *et al*. Detection of sub-shot-noise spatial correlation in high-gain parametric down conversion. *Phys Rev Lett* 2004; **93**: 243601.
- 25 Mosset A, Devaux F, Lantz E. Spatially noiseless optical amplification of images. *Phys Rev Lett* 2005; **94**: 223603.
- 26 Blanchet JL, Devaux F, Furfaro L, Lantz E. Measurement of sub-shot-noise correlations of spatial fluctuations in the photon-counting regime. *Phys Rev Lett* 2008; **101**: 233604.
- 27 Corzo NV, Marino AM, Jones KM, Lett PD. Noiseless optical amplifier operating on hundreds of spatial modes. *Phys Rev Lett* 2012; **109**: 043602.
- 28 Boyer V, Marino AM, Pooser RC, Lett PD. Entangled images from four-wave mixing. *Science* 2008; **321**: 544–547.
- 29 Brida G, Genovese M, Ruo-Berchera I. Experimental realization of sub-shot-noise quantum imaging. *Nat Photonics* 2010; **4**: 227–230.
- 30 Brambilla E, Caspani L, Jedrkiewicz O, Lugiato LA, Gatti A. High-sensitivity imaging with multi-mode twin beams. *Phys Rev A* 2008; **77**: 053807.

- 31 Brambilla E, Gatti A, Bache M, Lugiato LA. Simultaneous near-field and far-field spatial quantum correlations in the high-gain regime of parametric down-conversion. *Phys Rev A* 2004; **69**: 023802.
- 32 Agafonov IN, Chekhova MV, Leuchs G. Two-color bright squeezed vacuum. *Phys Rev A* 2010; **82**: 011801.
- 33 Davies E. *Computer and Machine Vision: Theory, Algorithms, Practicalities*. 4th edn. Waltham: Academic Press; 2012.
- 34 Brida G, Genovese M, Meda A, Ruo-Berchera I. Experimental quantum imaging exploiting multimode spatial correlation of twin beams. *Phys Rev A* 2011; **83**: 033811.
- 35 Cole R. Live cell imaging: the cell's perspective. *Cell Adh Migr* 2015; **8**: 452–459.
- 36 Pittman TB, Shih YH, Strekalov DV, Sergienko AV. Optical imaging by means of two-photon quantum entanglement. *Phys Rev A* 1995; **52**: R3429–R3432.
- 37 Meyers R, Deacon KS, Shih Y. Ghost-imaging experiment by measuring reflected photons. *Phys Rev A* 2008; **77**: 041801(R).
- 38 Ferri F, Magatti D, Gatti A, Bache M, Brambilla E *et al*. High-resolution ghost image and ghost diffraction experiments with thermal light. *Phys Rev Lett* 2010; **94**: 183602.
- 39 Brida G, Chekhova MV, Fornaro GA, Genovese M, Lopaeva ED *et al*. Systematic analysis of signal-to-noise ratio in bipartite ghost imaging with classical and quantum light. *Phys Rev A* 2011; **83**: 063807.
- 40 Gatti A, Brambilla E, Bache M, Lugiato LA. Ghost imaging with thermal light: comparing entanglement and classical correlation. *Phys Rev Lett* 2004; **93**: 093602.
- 41 O'Sullivan MN, Chan KW, Boyd RW. Comparison of the signal-to-noise characteristics of quantum versus thermal ghost imaging. *Phys Rev A* 2010; **82**: 053803.
- 42 Morris PA, Aspden RS, Bell JEC, Boyd RW, Padgett MJ. Imaging with a small number of photons. *Nat Commun* 2015; **6**: 5913.
- 43 Lopaeva ED, Ruo-Berchera I, Degiovanni IP, Olivares S, Brida G *et al*. Experimental realization of quantum illumination. *Phys Rev Lett* 2013; **110**: 153603.
- 44 Phan NM, Cheng MF, Bessarab DA, Krivitsky LA. Interaction of fixed number of photons with retinal rod cells. *Phys Rev Lett* 2014; **112**: 213601.
- 45 Avella A, Ruo-Berchera I, Degiovanni IP, Brida G, Genovese M. Absolute calibration of an EMCCD camera by quantum correlation, linking photon counting to the analog regime. *Opt Lett* 2016; **41**: 1841–1844.
- 46 Meda A, Ruo-Berchera I, Degiovanni IP, Brida G, Rastello ML *et al*. Absolute calibration of a charge-coupled device camera with twin beams. *Appl Phys Lett* 2014; **105**: 101113.
- 47 Peřina J, Haderka O, Allevi A, Bondani M. Absolute calibration of photon-number-resolving detectors with an analog output using twin beams. *Appl Phys Lett* 2014; **104**: 041113.
- 48 Agafonov IN, Chekhova MV, Iskhakov TS, Penin AN, Rytikov GO *et al*. Absolute calibration of photodetectors: photocurrent multiplication versus photocurrent subtraction. *Opt Lett* 2011; **36**: 1329–1331.
- 49 Bondani M, Allevi A, Zambra G, Paris MGA, Andreoni A. Sub-shot-noise photon-number correlation in a mesoscopic twin beam of light. *Phys Rev A* 2007; **76**: 013833.
- 50 Perina J Jr, Hamar M, Michálek V, Haderka O. Photon-number distributions of twin beams generated in spontaneous parametric down-conversion and measured by an intensified CCD camera. *Phys Rev A* 2012; **85**: 023816.
- 51 Iskhakov TS, Usenko VC, Andersen UL, Filip R, Chekhova MV *et al*. Heralded source of bright multi-mode mesoscopic sub-Poissonian light. *Opt Lett* 2016; **41**: 2149–2152.
- 52 Iskhakov TS, Chekhova MV, Leuchs G. Generation and direct detection of broadband mesoscopic polarization-squeezed vacuum. *Phys Rev Lett* 2009; **102**: 183602.
- 53 Tapster PR, Seward SF, Rarity JG. Sub-shot-noise measurement of modulated absorption using parametric down-conversion. *Phys Rev A* 1991; **44**: 3266–3269.
- 54 Whittaker R, Erven C, Neville A, Berry M, O'Brien JL *et al*. Absorption spectroscopy at the ultimate quantum limit from single-photon states. *New J Phys* 2017; **19**: 023013.
- 55 Simon DS, Jaeger G, Sergienko AV. Quantum information in communication and imaging. *Int J Quantum Inform* 2014; **12**: 1430004.



This work is licensed under a Creative Commons Attribution-NonCommercial-NoDerivs 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by-nc-nd/4.0/>

© The Author(s) 2017

Supplementary Information for this article can be found on the *Light: Science & Applications*' website (<http://www.nature.com/lisa>).