



ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

The 2021 IAEA software intercomparison for k(0)-INAA

Original

The 2021 IAEA software intercomparison for k(0)-INAA / Blaauw, M; D'Agostino, G; Di Luzio, M; Dung, Hm; Jacimovic, R; Dias, Mds; Semmler, R; Van Sluijs, R; Barradas, Np. - In: JOURNAL OF RADIOANALYTICAL AND NUCLEAR CHEMISTRY. - ISSN 0236-5731. - (2022). [10.1007/s10967-022-08626-1]

Availability:

This version is available at: 11696/75659 since: 2023-02-14T12:35:36Z

Publisher:

SPRINGER

Published

DOI:10.1007/s10967-022-08626-1

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)



The 2021 IAEA software intercomparison for k_0 -INAA

Menno Blaauw¹ · Giancarlo D'Agostino² · Marco di Luzio² · Ho Manh Dung³ · Radojko Jacimovic⁴ · Mauro Da Silva Dias⁵ · Renato Semmler⁵ · Robbert van Sluijs⁶ · Nuno Pessoa Barradas⁷

Received: 7 July 2022 / Accepted: 12 October 2022
© The Author(s) 2022

Abstract

In order to establish the variation between results in mass fractions due to software implementation, as measured by the k_0 -method for INAA, the IAEA has organized a software intercomparison. A complete set of test spectra and associated information was assembled. Efficiency curves, neutron spectrum parameters, correction factors and mass fractions were calculated with the participating programs (k0-IPEN, k0-INRIM, k0-DALAT, k0-IAEA and KayWin) using identical peak areas. In this paper, we report on the observed discrepancies, causes, remedies and future software developments. The test data, as well as intermediate results and observed mass fractions of the certified reference material BCR-320R “channel sediment” are available through the IAEA on request. The variations in concentrations attributed to differences between the programs were initially found to be 5.6 and 7.9%, for certified and uncertified concentrations, respectively. After the certified concentrations had been made available to the participants and they had been allowed to improve their programs, the variations found were 2.7 and 3.4%, respectively. The main identified remaining causes of variation are differences in the procedures used for detector efficiency characterisation and neutron spectrum parameter determination.

Keywords Neutron activation analysis · k_0 -method · Software intercomparison · TRIGA reactor · Neutron flux parameters · Germanium detector

Introduction

The k_0 -method for INAA

The k_0 -standardization method of Neutron Activation Analysis (k_0 -NAA) was launched in 1975 by Simonits et al. [1]. The method is based on experimentally determined k_0 -factors, which are independent of irradiation and measuring conditions. Due to well-known characteristics of gold and its applicability in NAA, the element is chosen as the standard for k_0 -NAA, and all necessary nuclear data of investigating nuclides are normalized to the gold nuclear data. The method has been continuously improving, including its nuclear database, the so-called k_0 -database [2]. Due to its wide applicability and the possibility of panoramic analysis, k_0 -NAA has been introduced in many laboratories around the world. Consequently, some laboratories developed their own programs (e.g. k0-IPEN, k0-INRIM [3], k0-DALAT) besides a few that are used in several locations (i.e. KayWin [5–8] and k0-IAEA [9, 10]).

✉ Menno Blaauw
m.blaauw@tudelft.nl

- ¹ Reactor Institute Delft, Delft University of Technology, Mekelweg 15, 2629 JB Delft, The Netherlands
- ² Istituto Nazionale di Ricerca Metrologica (INRIM), Strada delle Cacce 91, 10135 Turin, TO, Italy
- ³ Center for Nuclear Technologies (CNT), 217 Nguyen Trai Street, District 1, Ho Chi Minh City, Vietnam
- ⁴ Jozef Stefan Institute, Ljubljana, Slovenia
- ⁵ Research Reactor Center (CERPq), Instituto de Pesquisas Energéticas e Nucleares (IPEN-CNEN/SP), Av. Prof. Lineu Prestes, São Paulo, SP 2242, Brazil
- ⁶ k0-Ware, Schaesbergerweg 25, 6415 AB Heerlen, The Netherlands
- ⁷ Vienna International Centre, International Atomic Energy Agency, PO Box 100, 1400 Vienna, Austria

Sources of uncertainty in the k_0 -method

Various papers have been written on the estimation of the combined standard uncertainty in mass fractions as obtained with the k_0 -method [11]. It has been concluded that the stated uncertainties in the k_0 -factors, typically in the order of 1% for the main peak energies, only represent a minor part of the overall uncertainties, typically being about 3.5%. Uncertainty propagation taking all contributions into account is challenging because of strong correlations between the nuclear data (i.e. half-life, Q_0 , effective resonance energy) and the k_0 -factors, a strong correlation between the neutron spectrum parameters f and α , as well as strong correlations between mass fractions obtained through different gamma-ray energies or different measurements of the same sample. The only independent source of uncertainty for individual mass fractions to be taken into account is the uncertainty in the peak areas used due to counting statistics. For that reason, it is relatively easy to propagate, and sometimes the only source that is taken into account. Some other sources of uncertainty that affect all mass fractions equally, e.g. the sample mass or the measured flux, can be also easily propagated.

Previously, the International Atomic Energy Agency (IAEA) conducted intercomparisons for gamma-ray spectrometry [12, 13], alpha-spectrometry [14, 15], particle-induced X-ray emission (PIXE) programs [16], low-level gamma-ray spectrometry [17, 18], ion-beam analysis (IBA) techniques including Rutherford backscattering (RBS), elastic-recoil detection analysis (ERDA) and nuclear reaction analysis (NRA) programs [19], and Particle Induced Gamma-ray Emission (PIGE) [20]. Even though it was concluded in these intercomparisons that the programs were appropriate for their stated purpose in general, significant differences between programs were found in some cases. In any case, it is important to ascertain and quantify the influence of using different programs on the results obtained. A limited intercomparison was done for two k_0 -NAA programs [21]. While it was concluded that “in general, a good agreement was found between the elements contents determined by both programs and certified, non-certified and/or assigned values” of the certified reference materials analysed, the authors of the programs were not involved in the intercomparison, and therefore there was no guarantee that the programs were used optimally. Also, several other k_0 -NAA programs exist or were developed since, which were not included in the intercomparison. Therefore, it was decided to perform the current intercomparison. All known k_0 -NAA programs were invited to participate, and a call to participation was issued through the k_0 -International Scientific Committee. In the end, out of six programs identified, five participated in the exercise.

Experimental

Selection of participants

Existing programs were located through the k_0 -users community. Three programs were in use only in the institute where they had been developed (k0-IPEN, k0-INRIM and k0-DALAT), another two in multiple locations (KayWin [5–7], k0-IAEA [9, 10]). The KayWin program is based on the original source code developed by De Corte et al., the others were developed independently. In each case, the developer of the program was invited to participate.

Description of the participating programs

k0-DALAT

The k0-DALAT (2017 version) program is capable of automatic gamma-ray spectrum processing, the calculation of the detector efficiency, the calculation of neutron flux parameters and the calculation of mass fraction, uncertainty and detection limit of the element of interest. For the full-energy-peak detector efficiency, a 4th-order polynomial is used.

The Angle [22] program (Version 3) is used to convert the reference efficiency to actual counting geometries. The TrueCoinc program [23] is used to compute true-coincidence summing correction factors.

For the neutron spectrum parameters f , α and the mass fractions, the program can use two options, i.e. “triple bare” or “Cd-covered”, for thermal and epithermal INAA, respectively, using Al-Au plus Zr. If Ni and Lu are present, the fast flux and neutron temperature are determined as well.

k0-IAEA

With the k0-IAEA (version 9.10, 2021) program, the peak-to-total efficiency curve is determined using only a ^{137}Cs spectrum. Afterwards, the reference full-energy-peak efficiency curve can be determined using only a ^{152}Eu spectrum measured at any distance or even inside a well-type detector. A restricted sixth-order polynomial as developed by Gunnink [24] is fitted to the measured peak areas to this end, taking true-coincidence summing corrections into account in the process. To convert the reference efficiency curve to other counting geometries, if needed, Monte Carlo calculations, that are embedded in k0-IAEA, are employed for a predefined set of gamma-ray energies, covering the range from 50 to 3000 keV, to a user-defined

statistical precision. Sample geometry and matrix composition are taken into account in these simulations, as well as all available information on the detector, absorbers and sample containers.

k0-IAEA can use any flux monitor composition with enough activation products to determine the neutron spectrum parameters, using the multiple-bare method.

k0-INRIM

The k0-INRIM (version 2.0, 2021) program was firstly developed to obtain results with uncertainty budgets as prescribed by the Guide to the Expression of Uncertainty in Measurement (GUM) taking into account correlations and based on the spreadsheet technique [25]. This requires the adoption of an original approach for the detector efficiency characterisation based on a fully experimental procedure without the need for Monte Carlo simulations [26].

Built-in functions are implemented for the automatic determination of (i) monitor to analyte efficiency ratio, (ii) neutron flux parameters by “triple bare method”, (iii) neutron flux gradient at the irradiation channel and (iv) peak-to-total ratios. In addition, a number of corrections are evaluated and applied, e.g. (i) sample positioning misplacements, (ii) extended sample geometry, (iii) gamma self-absorption, (iv) true-coincidence, (v) counting pile-up, (vi) blank and (vii) moisture.

k0-IPEN

The k0-IPEN (version 1.7, 2021) program is able to get full-energy peak efficiency curves for all counting positions based on acquired gamma spectra. The peak-to-total curves are obtained using spectra of mono-energetic sources. A fourth-order polynomial is fitted to the available data at the reference position. The true-coincidence summing correction is calculated by a routine imbedded inside the package, applying the Semkow matrix methodology [27].

For the actual counting geometries, the efficiencies are determined by Monte Carlo calculations using MCNP6 [28] and verified using spectra measured at those distances.

The neutron spectrum parameters can be determined by the “triple bare” method or the “Cd-cover” method—“triple bare” was used in the intercomparison.

KayWin

Kayzero for Windows [5] (a.k.a. “KayWin” V3.37) has built-in options for neutron spectrum calibration, i.e. the “multi monitor bare”, “Cd-cover” and “Cd-ratio” methods as well as the “triple bare” and the “special Au-Zr triple bare” methods.

The efficiency conversion from one counting geometry to another [7] is performed using effective solid angles calculated using SOLANG, coincidence correction factors are calculated using COINCALC, both based on the original programs from Frans De Corte [29]. Because these programs are based on fits of the reference efficiency using three or four regions with 2nd or 3rd-order polynomials, this is still the procedure in the present program. The reference full-energy-peak efficiency was fitted using data of all measurements of calibration standards provided. Fine-tuning of the dead-layer thickness and vacuum-gap can be performed automatically using true-coincidence-free sources if needed.

For the determination of the peak-to-total efficiency curve, all given mono-energetic spectra are used. All possible corrections are implemented: nuclide identification, gamma interference [6] correction, interference due to threshold reactions and fission products, burn-up and more.

Intercomparison methodology

It was decided to minimize all sources of variation from the intercomparison, except for the differences between the participating programs. To that end, the participants were supplied with the same experimental data, i.e. for efficiency calibration, irradiation facility characterization, sample spectra et cetera, including the peak areas and energies. The data provided allowed for different approaches to detector and irradiation facility characterization. One co-author (Jacimovic) was invited by the IAEA to prepare the test spectra, select the elements for the participants to report on, determine the peak areas, distribute the data and gather the first set of results, in order to prevent prior knowledge from affecting the outcome. The elements and peaks to be reported on by the participants were selected taking into account that induced radionuclides could be typically detected in a sediment sample via short or long irradiation including a wide energy interval of gamma rays (from 100 to 3000 keV). High counting statistics for the peaks involved were required, in order to allow for testing the propagation of other sources of uncertainty to the results. Certified mass fractions of some elements in the sample were required in order to check the accuracy of the data reported. Last but not least, elements were selected to address known complexities of the k_0 -method, such as the necessity to correct for coincidence summing, or to correct for spectral and primary interferences. This co-author did not participate as a software user. The participants only reported on the preselected elements and peaks.

The results from the participants would have turned out to be identical, if all had implemented and used the k_0 -method in exactly the same way, i.e. using identical irradiation facility characterization methods, efficiency curve modelling and efficiency conversion methods, all as described in

De Corte's thesis [29]. The uncertainties as reported by the programs, taking counting statistics as well as other sources of uncertainty into account, are less relevant in this between-programs variation respect, because most such sources of variation were eliminated in the experimental design.

At the same time, the fact that a certified reference material was used allowed for a comparison between the results from individual programs and the certified values, pointing to possible improvements where the results showed differences. The reported uncertainties are highly relevant in this respect.

Spectra for detector characterization

All spectra were acquired with a 40% relative-efficiency germanium detector in a vertical dipstick configuration. Point sources as shown in Table 1 were counted at five well-defined platform positions at distances between point source and end cap of the detector of 203.4, 123.4, 83.4, 43.4 and 23.4 mm – the ring-shaped point-source holders introduced the extra distance of 1.4 mm. Four sources were measured only at the 203.4 mm reference position. Counting times were sufficient to expect counting-statistics uncertainties of less than 0.1% for the relevant, main peaks of each nuclide.

Spectra for irradiation facility characterization

To allow for irradiation facility characterization, i.e. the determination of the k_0 -parameters f and α , metal foils of pure Zn, Fe, Zr and a IRMM-530R Al-Au alloy were stacked, irradiated simultaneously during 45 min, in the IC-40 of the 250 kW TRIGA Mark II reactor of the Jozef Stefan Institute (JSI) and measured separately. Two irradiations were performed, one with 1 mm thick cadmium cover (Cd-box, diameter 1 cm, height 2 cm, mass ~ 5 g) and one without ("bare"). Measurements were performed after

waiting times ranging from about 8 h to three days. More detail is shown in Table 2.

Flux-monitor and sample spectra

Two samples were prepared by weighing aliquots of certified reference material BCR-320R into cylindrical polyethylene capsules with internal radius and height of 4 mm and 4 mm, respectively. As prescribed by the certificate, a moisture correction factor of 0.974 was determined by oven-drying about 1 g of not-to-be-analyzed material at 105 °C until constant mass was attained. One sample was irradiated in the TRIGA reactor of the Jozef Stefan Institute on top of a IRMM-530R Al-Au alloy flux monitor for 30 s. The other sample was irradiated for 12 h in the same position, but this time in a stack with three additional samples and two additional flux monitors on top. Sample and flux-monitor details, irradiation times and counting positions are provided in Table 3.

A blank was not included. Instead, information on the blank composition was provided to all participants – only affecting the results for Cr.

Determination of peak areas

Peak areas and energies were determined using Hyperlab 2014 [30] from the raw spectra and made available to the participants in various file formats, along with the raw spectra (that were to be used exclusively for detection limit estimation).

Software intercomparison—first round

All spectra and experimental data were provided to the participants, including indicative data on the matrix composition, but without revealing the BCR-320R identity or the fact that both samples were of the same material. The participants filled out extensive spreadsheets with information

Table 1 Calibration sources and distances from the detector

Nuclide	Producer	Ref. date	Activity (Bq)	Unc.% (k = 2)	Distances
Am-241	CMI, Czech	20 Oct2014	2.008E + 04	0.8	all five
Am-241	LMRI, France	15 June1988	3.790E + 04	2.0	all five
Cd-109	CMI, Czech	20 Oct2014	4.140E + 04	2.0	all five
Co-57	CMI, Czech	20 Oct 2014	5.008E + 04	0.8	all five
Mn-54	CMI, Czech	20 Oct 2014	4.004E + 04	2.0	all five
Zn-65	CMI, Czech	20 Oct 2014	3.753E + 04	1.2	all five
Cs-137	LMRI, France	15 June 1988	3.972E + 04	1.5	all five
Ba-133	PTB, Germany	01 Sept 1989	2.360E + 05	1.5	203.4 mm only
Eu-152	PTB, Germany	01 Sept 1989	2.160E + 05	2.0	203.4 mm only
Ra-226	PTB, Germany	01 Sept 1989	1.587E + 05	2.81	203.4 mm only
Co-60	CMI, Czech	20 Oct 2014	2.319E + 04	1.2	203.4 mm only
Au-198	JSI**		Unknown	Unknown	all five

** Source prepared by neutron activation of Al-0.1% Au alloy on 30/06/2016 at 17:35:00.

Table 2 Flux monitor foils used for irradiation facility characterization

Foil material	Purity (%)	Radius (mm)	Height (mm)	Mass (“bare” irradiation) (g)	Mass (Cd-covered irradiation) (g)	Sample-detector end cap distances (mm)
Zn	99.99 +	3.5	0.025	0.00682	0.00693	122 and 22
Fe	≥ 99.996	3.2	0.1	0.02475	0.02196	22
Zr	99.8	3.0	0.14	0.02542	0.02507	202 and 22
Al-0.1% Au	0.1003% Au	3.4	0.1	0.00934	0.00933	202

Table 3 Sample and flux monitor details

Sample	Radius (mm)	Height (mm)	Measured mass (g)	Dry mass used (g)	Irradiation time	Sample-detector distances and approximate waiting times
BCR-320R	4.0	4.0	0.20766	0.20234	30 s	after 8 m at 202 mm, after 39 min at 202 mm
Al-0.1%Au	3.4	0.1	0.00951	9.539 µg of Au	30 s	after 6 h at 82 mm
BCR-320R	4.0	4.0	0.19343	0.18847	12 h	after 4 d at 202 mm, after 9 and 27 d at 22 mm
Al-0.1%Au	3.4	0.1	0.00933	9.358 µg of Au	12 h	after 6 d at 202 mm
Al-0.1%Au	3.4	0.1	0.00915	9.177 µg of Au	12 h	after 6 d at 202 mm
Al-0.1%Au	3.4	0.1	0.00930	9.328 µg of Au	12 h	after 6 d at 202 mm

Table 4 Improvements made to the participating programs

Name	Improvement made
KayWin	None
k0-DALAT	Fine-tuning efficiency curve (4th-order) at high energies beyond 1408 keV was performed, using the Na-24 peaks.; The Hg-Se interference at 279 keV was corrected for; Peak areas as distributed to all participants were now used;
k0-IAEA	The method for peak-to-total detector characterization was improved (it now includes the peak area in the “total”); The k_0 -data used were adjusted to the data in the k_0 database, resulting in minor changes in a few k_0 -constants and half-lives
k0-INRIM	The measurement equation was improved with parameters modelling i) the gamma self-absorption correction, ii) gamma interference correction, iii) moisture correction and iv) the non-linearity of gamma efficiency variation versus detector distance; Interference correction was performed manually by the user; Improvement of the neutron flux characterisation with the quantification of the fast neutron flux
k0-IPEN	Correction of spectral interferences was implemented; The nuclear database was adjusted according to the k_0 database; Threshold reactions interferences were corrected for manually; Values for f and α were adjusted due to correction of the efficiencies for the reference position

on the nuclear data and k_0 -constants used, results obtained from single peaks, from all peaks within a spectrum and from all spectra combined. All results were submitted to the non-participating co-author. Two outliers were spotted and corrected before processing the first-round data: One participant had failed to correct for the $^{27}\text{Al}(n,p)^{27}\text{Mg}$ reaction contribution, and another had made a typo when entering the 12 h irradiation time for the flux monitors. These mistakes were considered “out of scope” for the intercomparison and corrected before comparing the results. One participant (k0-DALAT) did not use the peak areas that were to be used by the others and determined them within the program, as this is the normal way to operate this program.

Software intercomparison—second round

The certificate of the reference material was made available to all participants along with all results obtained. All were asked to investigate the causes of differences between the results for the certified elements, and to improve their program or its usage if deemed necessary. Intermediate results on efficiency curves were shared and compared. All were asked to resubmit their results, with more detail on intermediate computational steps than in the first round.

The improvements were made to the various programs are shown in Table 4.

Efficiency curves

In the k_0 -method, relative efficiencies matter most, i.e. ratios of detection efficiencies for sample and flux monitor peak energies. The participants reported absolute efficiencies for all peak energies used, for the reference counting geometry as well as for actual sample and flux monitor counting geometries. From these, the ratios of efficiencies as employed in concentration calculations were calculated to allow for comparison between participants. For the mass fraction calculations, the efficiency ratios for the sample at 202 mm versus the flux monitor at 82 mm from the end cap are relevant for the short-irradiation experiment. For the long-irradiation experiment, the ratios of the sample at 202 and 22 mm versus the flux monitor at 202 mm from the end cap are relevant.

Most participants converted detection efficiencies from the reference counting geometry to the actual sample counting geometry by computational methods, as described above.

k0-DALAT fitted its 4th-order polynomial to the available datapoints.

In k0-IAEA, the peak-to-total calibration was obtained from the ^{137}Cs source measured at 23.4 mm from the end cap, and the full-energy-peak efficiency curve from the ^{152}Eu spectrum measured at 203.4 mm. The Monte Carlo efficiency-conversion simulations were carried out to a 0.1% statistical precision.

With the k0-INRIM program a 6th-order polynomial was fitted to the datapoints for the reference position, using all gamma lines of all sources. For non-reference counting geometries, curves were obtained by calculating measured count-rate ratios of coincidence-free gamma lines from sources with respect to the reference position; in those positions the count rate ratio data were fitted with similar exponential polynomial equations. In case the non-reference

counting geometries did not coincide with the actual sample counting geometries, vertical efficiency corrections were performed by modelling the typical inverse square trend, also evaluated from count rate ratio data.

The k0-IPEN program used all available spectra and datapoints, from 53 up to 2000 keV, for the full-energy-peak efficiency calibration at the reference position. For the 2754 keV line of ^{24}Na however, a Monte Carlo method was used. For the peak-to-total efficiency curve, all mono-energetic spectra were used.

For efficiency conversion, KayWin uses SOLCOI [31], which is a combination of SOLANGE [32] and the coincidence correction software COINCALC by De Corte and De Wispelaere.

Relative efficiencies determine the mass fractions primarily, but the absolute efficiencies affect the mass fractions a little through the corrections for true coincidence summing. To compare the programs in this respect, coincidence correction factors for all peaks were tabulated.

Neutron flux parameters

All programs except KayWin used the “triple bare” method for the determination of f and α . KayWin used the Cd-ratio method. The fast flux was determined by all participants using the $^{27}\text{Al}(n,\alpha)^{24}\text{Na}$ and/or the $^{54}\text{Fe}(n,p)^{54}\text{Mn}$ reactions.

Results

Software intercomparison first round

For each investigated element, the ratio of the single reported mass fraction to the unweighted average of all reported mass fractions was calculated. The results are

Fig. 1 Mass fractions as reported by the programs divided by the average of all programs for that element. Elements As through Zn were certified, Al through Sm were not





Fig. 2 Mass fractions as reported by the programs divided by the certified value for that element. The error bars indicate the $k=1$ confidence interval for the certified values



shown in Fig. 1. The variation due to program differences is computed as the standard deviation of these ratios. The observed overall variation is 6.4%, while the observed variation for certified and uncertified elements is 5.6 and 7.9%

For each certified element, the ratio of the reported mass fraction to the certified value was calculated. The resulting ratios are shown in Fig. 2.

In addition, the overall quality of the results of each program was assessed with a Q^2 -score, defined as

$$Q_j^2 = \frac{1}{N} \sum_{i=1}^N \frac{(c_{ij} - c_{ci})^2}{s_{ci}^2} \quad (1)$$

where N is the number of mass fractions, c_{ij} is the mass fraction reported for element i by program j , c_{ci} is the certified mass fraction and s_{ci} is the 1 standard deviation uncertainty in the certified mass fraction.

The lower the Q^2 -score, the better the overall agreement with the certified values. Outliers heavily affect the value and are not removed because they are symptomatic of possible systematic errors. A score below 1 should not be achievable, normally, since any analysis method used would be expected to introduce uncertainties of its own that are not accounted for in the Q^2 -calculation.

The statistical control of each program was expressed in a reduced χ_r^2 -score, defined as

$$\chi_{rj}^2 = \frac{1}{N} \sum_{i=1}^N \zeta_{ij}^2 \quad (2)$$

where the ζ_{ij} are the zeta-scores, defined by

$$\zeta_{ij} = \frac{c_{ij} - c_{ci}}{\sqrt{s_{ij}^2 + s_{ci}^2}} \quad (3)$$

where s_{ij} is the 1 standard deviation uncertainty for concentration i as reported by program j .

The zeta-scores indicate the agreement of the result with the certified values, taking the uncertainty reported by the program into account, and are shown in Fig. 3.

At 11 degrees of freedom, the χ_r^2 -scores should lie between 0.35 and 2 for a program's reported uncertainties to properly represent the accuracy of the reported mass fractions with 95% confidence: a score below the lower limit or above the upper limit indicates a likely systematic overestimation or underestimation of the combined uncertainty evaluated by the program, respectively.

The resulting Q^2 - and χ_r^2 -scores are presented in Table 5.


Software intercomparison second round

As in the first round, the ratio of the single reported mass fraction to the unweighted average of all reported mass fractions was calculated. The resulting ratios are shown in Fig. 4. The observed variation for all elements is 3.0%, while for the certified and uncertified elements, it is 2.7 and 3.4%, respectively.

The ratios of the reported mass fractions to the certified values were calculated. The resulting 2nd round ratios are shown in Fig. 5.

The zeta-scores for the 2nd round are shown in Fig. 6.

The 2nd-round Q^2 -scores and χ_r^2 -scores are presented in Table 6.

Fig. 3 Zeta-scores for the certified elements in the first round**Table 5** Q^2 - and χ_r^2 -scores for the participating programs

program name	Q^2	χ_r^2
k0-INRIM	2.91	1.35
k0-IAEA	1.45	0.69
k0-IPEN	3.22	2.07
k0-DALAT	2.43	0.41
KayWin	0.55	0.37

each such ratio was divided by the average ratio over all participants, and the ratio-of-ratio was calculated for the sample at 202 mm with the flux monitor at 82 mm, for both sample and monitor at 202 mm, and for the sample at 22 mm with the flux monitor at 202 mm. The latter are shown in Fig. 7 to exemplify that the variations differ from peak to peak.

The variation in these ratios of ratios amounts to 2.1, 1.9 and 3.1% for the three cases, respectively.

True coincidence summing correction factors for selected peaks are shown in Fig. 8. The overall variation observed after diving each correction factor by the average for that peak energy was found to be 0.4%.

Efficiency curves and true coincidence summing

For a given pair of sample and flux monitor gamma-ray peaks, the mass fraction is proportional to the inverse ratio of the two corresponding detection efficiencies. The value of


Fig. 4 Mass fractions as reported by the programs in the second round, divided by the average of all programs for that element. Elements As through Zn were certified, Al through Sm were not

Fig. 5 Mass fractions as reported by the programs divided by the certified value for that element. The error bars indicate the $k=1$ uncertainty in the certified values




Fig. 6 Zeta-scores for the certified elements in the second round




Table 6 Q^2 - and χ_r^2 -scores for the participating programs in the second round

program name	Q^2	χ_r^2
k0-INRIM	1.13	0.49
k0-IAEA	0.66	0.52
k0-IPEN	0.86	0.46
k0-DALAT	1.10	0.51
KayWin	0.49	0.37

Neutron spectrum parameters

All participants used the efficiency curves as determined and the spectra provided to determine the conventional subcadmium flux ϕ_0 , f and α as well as the fast neutron flux for the irradiation position used. The 2nd-round results are shown in Table 7.

The thermal and epithermal neutron self-shielding factors as applied in the f and α determinations, being a possible for source of discrepancy, were reported by the participants and are shown in Table 8.





Fig. 7 Program-specific efficiency ratios, divided by the average over all programs, for sample peaks measured at 22 mm versus the flux monitor 411 keV peak measured at 202 mm

Fig. 8 True coincidence summing correction factors as reported by the participants, for the nuclides, peaks and counting geometry where coincidence summing has the most extreme effect (i.e. for the 1173 and 1332 keV peaks of ^{60}Co , 603 and 1691 keV peaks of ^{124}Sb , and the 889 and 1120 keV peaks of ^{46}Sc , at 22 mm from the end cap of the detector)



Discussion

Mass fractions

Reasonably good agreement is observed between the mass fractions as reported by the participating programs. The

observed variation between them appears to be due mostly to differences in efficiency curves and the resulting efficiency ratios used in neutron spectrum parameter and mass fraction calculations. The procedure selected for neutron spectrum parameter determination (e.g. bare versus Cd-cover methods) also contributes.

Table 7 Neutron spectrum parameters as reported. More significant digits than usual are shown because all programs used the same input data, but propagated counting statistics and other sources of uncertainty

program	method	ϕ_0 subcadmium flux $\text{cm}^{-2} \text{s}^{-1}$	unc $\text{cm}^{-2} \text{s}^{-1}$	f	unc	α	unc	fast flux $\text{cm}^{-2} \text{s}^{-1}$	unc $\text{cm}^{-2} \text{s}^{-1}$
k0-DALAT	bare, Fe	1.067E+12	7.4E+10	26.50	5.30	0.012	0.005	1.910E+11	9.5E+09
k0-IAEA	bare, Fe	1.092E+12	4.0E+09	26.49	1.05	0.017	0.007	1.369E+11	3.0E+09
k0-INRIM	bare, Fe	1.120E+12	3.7E+10	30.25	2.61	−0.006	0.014	1.293E+11	4.2E+09
k0-IPEN	bare, Al	1.093E+12	1.5E+10	29.06	1.30	−0.009	0.004	1.410E+11	1.9E+10
KayWin	Cd-ratio, Fe	1.072E+12	1.0E+10	26.80	0.80	0.000	0.010	1.420E+11	1.4E+09
KayWin*	bare, Fe			30.46	n/a	−0.009	n/a		

Table 8 thermal and epithermal corrections as applied by the participants. Q_0 -values are shown to indicate the relevance of the epithermal self-shielding correction, which increases with Q_0

	Q_0	k0-DALAT		k0-IAEA		k0-INRIM		k0-IPEN		KayWin	
		g_{th}	g_{epi}	g_{th}	g_{epi}	g_{th}	g_{epi}	g_{th}	g_{epi}	g_{th}	g_{epi}
Au-197→Au-198	15.7	1.000	1.000	1.000	0.998	1.000	1.000	0.999	0.998	1.000	0.996
Zr-94→Zr-95	5	1.000	1.000	1.000	0.983	1.000	0.981	1.000	0.975	1.000	0.988
Zr-96→Zr-97	248	1.000	1.000	1.000	1.000	1.000	0.970	1.000	0.970	1.000	0.966

From the first-round results, it is clear that Hg and Mg are the most troublesome elements (Fig. 1). The ^{203}Hg -peak at 279 keV suffers from a strong interference with the ^{75}Se peak at the same energy, the latter being subject to strong coincidence summing effects. Mg suffers from a strong contribution from the $^{27}\text{Al}(n,p)^{27}\text{Mg}$ reaction. All the other elements were relatively straightforward, i.e. essentially free from any kind of interference.

The Q^2 -score of 0.55 for the KayWin results in the first round, followed by similar values for KayWin and k₀-IAEA in the second round, are well below unity. This finding suggests overestimation of the certified mass fraction uncertainties, or it might be due to the fact that the k₀-method was employed in certifying the reference material to begin with, resulting in a correlation between the certified mass fractions and the mass fractions found in the current intercomparison. All the χ_r^2 -values are lower than the Q^2 -values, which is due to the inclusion of the uncertainty estimates of the participating programs themselves. All programs exhibit good statistical control in both rounds at the 95% confidence level.

These observations imply that the current intercomparison does not reflect on the inherent uncertainties associated with k₀-INAA results: the reference material was not certified with uncertainties small enough to shed light on this issue.

The observation that the results in the second round are much better than in the first round, for most participating programs, indicates that there was room for improvement. Some improvements impacted many results, others only individual mass fractions. All improvements were made after sharing all results and knowing the desired,

tainty to the results (*KayWin bare data for information only, not used in other computations)

certified outcome, as well as after fruitful discussions among participants.

It is unknown how these improvements would affect results obtained for materials with unknown composition. Therefore, the whole intercomparison would need to be repeated with a newly selected, unknown reference material in order to yield a meaningful value for the variation in k₀-NAA results due to implementation differences between programs. Alternatively, elements not considered in the intercomparison, but clearly present in the gamma-ray spectra, could be employed to this end – but in the absence of certified mass fractions, a reduction in observed variation would not necessarily demonstrate better accuracy.


Neutron spectrum parameters

There is no clear relation between the different results for f and α versus the different self-shielding corrections as applied.

In order to assess the impact of the observed differences in reported f and α -values, the neutron flux component of the standard k₀-NAA mass-fraction calculation, given by $(f + Q_0(\alpha))_m / (f + Q_0(\alpha))_a$ (subscripts m and a refer to monitor and analyte elements), was calculated for each relevant activation reaction for each pair of f , α -values, since this is the ratio that accounts for the neutron spectrum influence on the quantified mass fraction calculations.

The results are shown in the Fig. 9. That is to say, the neutron flux component values impact the reported mass fractions by the factors shown. For example, the smallest

Fig. 9 Deviation-from-average factors due to differences in neutron flux parameters. Lines were drawn only to guide the eye



impact in terms of relative variation is for Sm (about 0.2%), and the highest is for U (4.3%).

The factors shown in Fig. 9 were applied to the original reported mass fractions, thus eliminating the impact of variation in the f and α parameters, and resulting in new overall variations in mass fractions of 3.0% overall, 2.7% for the certified elements and 3.5% for the uncertified elements, so no major overall improvement was observed. Nevertheless, relative differences of the order of a few percent between k0-INRIM, k0-IPEN and k0-IAEA, k0-Dalat, KayWin were clarified.

Efficiency curves

Efficiency ratios, such as the ones compared in Fig. 7, are used directly in k_0 mass fraction calculations, where sample-to-monitor efficiency ratios are inversely proportional to calculated mass fractions. The differences between these, for individual programs, necessarily correspond to differences between mass fraction results. The relations between such mass fraction differences and the efficiency ratios are straightforward when considered for a single gamma-ray peak, but become complex when multiple peaks from multiple spectra are involved in a mass fraction calculation, as was the case for most participating programs.

It is obvious from the data in the figure mentioned, that mass fraction differences of several% are to be expected from the observed differences in efficiency ratios – of up to 17% for mass fractions that would have been obtained from the 2754 keV peak of Na as measured at 22 mm from the end cap (Fig. 7), if the 1368 keV would have been absent or disregarded.

Therefore, determination of the detector's efficiency curves, and conversion of these from one counting geometry to another, clearly is a major source of variation in mass fractions between programs in the current intercomparison.

Gamma-ray interferences

Gamma-ray interference corrections are applied by the different participants, in the second round, but played a minor role in the determination of virtually all element mass fractions, except for the impact of Se on Hg, as became obvious in going from the first-round to second-round results. All participants except k0-IAEA performed the interference correction the same way, that is, by calculating the Se mass fraction from the interference-free Se peaks, and deducting the Se contribution from the Hg peak area at 279 keV afterwards. The k0-IAEA program solves all interferences by considering all peak areas to be the linear sum of the contributions from the elements, and solving the resulting system of linear equations by least-squares methods. The difference between the first and second round results for k0-IAEA was due to imperfect coincidence summing corrections due to imperfect peak-to-total determination, as mentioned above.

Threshold reaction interferences

k0-INRIM, k0-IPEN and KayWin do not apply threshold interference reactions automatically. The first two require the user to perform the calculations. KayWin leaves the choice to apply a predetermined correction by the program to the user. k0-IAEA always corrects automatically. k0-DALAT will automatically correct if a spectrum of threshold monitor (i.e. Ni) is selected or ratio of thermal to fast neutron flux (f_T)

is entered. Actually, this exercise used Fe monitor instead of Ni (which is formulated in the program), so the correction was not performed in 1st round, but it was done in the 2nd by inputting the value f_{f} .

The capture cross sections used by the different programs, not having been standardized in the k_0 -library, contribute an estimated 7.4% to the variation in results for Mg, as well as in the reported values for the fast neutron flux as shown in Table 7.

Conclusions

An intercomparison exercise of five k_0 -NAA software packages was performed, using experimental data taken from a certified reference material obtained with a well-characterized system. Reasonably good agreement between programs is observed. A moderate amount of variation is present that must be due to differences between the participant programs, since experimental sources of variation were eliminated from the intercomparison. For the certified elements in the second round, the agreement between the programs was quite good (2.7% variation over 11 elements). The larger variation (3.4% over 6 elements) for uncertified elements is mostly due to the $^{27}\text{Al}(\text{n,p})^{27}\text{Mg}$ threshold interference reaction. The ^{75}Se interference with Hg determination through the 279 keV peak of ^{203}Hg remaining uncorrected in some first-round results played a major part there.

At the same time, all programs exhibited good statistical control in the sense that reported mass fractions agreed with certified mass fractions to within the specified uncertainties. This also demonstrates the validity of the data in the k_0 -database to within the uncertainties reported and certified in this intercomparison.

Interference correction procedures are not integral to the k_0 -method, and require the user to perform them in some cases (e.g. k_0 -INRIM, k_0 -IPEN). Threshold reaction interferences rely on fast-neutron capture cross sections or other methods that are typically determined in-house for each irradiation facility.

We established that the detector characterization, the radionuclides and selection of peaks, as well as the polynomials used to that end, impact the final mass fractions to a noticeable degree. In addition, the procedure adopted to characterize the irradiation facility (“Cd-cover” versus “triple bare”) plays a part.

For the detector and irradiation facility characterization, choices have to be made related to what monitors and source to use, as well as to the method to use for f and α determination. These choices will also affect the results. Based on only one dataset and personal experience, different choices were made by the users in this intercomparison, resulting in differences between programs.

We recommend the following to achieve the best possible results with the k_0 -method:

- to pay very close attention to the detector characterization step, using the largest possible number and energy range of gamma-ray lines;
- to investigate possible errors affecting f and α determinations depending on the adopted method;
- and to extend the k_0 -method with respect to the fast flux and the threshold capture reactions, in order to minimize the required calibration efforts.

Declarations

Conflict of interest This paper does not incur potential conflicts of interest.

Human and animal rights No human test subjects or animals played a part.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

1. Simonits A, De Corte F, Hoste J (1975) Single-comparator methods in reactor neutron activation analysis. *J Radioanal Chem* 24:31–46
2. Jaćimović R, De Corte F, Kennedy G, Vermaercke P, Revay Z (2014) The 2012 recommended k_0 database. *J Radioanal Nucl Chem* 300:589–592
3. D'Agostino G, Di Luzio M, Oddone M (2020) The k_0 -INRIM software: a tool to compile uncertainty budgets in neutron activation analysis based on k_0 -standardisation. *Meas Sci Technol* 31:017002
4. Di Luzio M, D'Agostino G (2022) The k_0 -INRIM software version 2.0: presentation and an analysis vademecum *J Radioanal Nucl Chem* (accepted)
5. van Sluijs R, Bossus DAW (1999) Optimized data evaluation for k_0 -based NAA. *J Radioanal Nucl Chem* 239:601–603
6. Van Sluijs R, Bossus DAW, Vervecken D (1994) Commercial k_0 -based neutron activation analysis at DSM Research (The Netherlands). *Acta Physica Hungarica* 75:355–359
7. Bossus DAW, van Sluijs R (1998) The influence of sample properties and sample geometry on the accuracy of gamma-ray spectrometric analyses. *J Radioanal Nucl Chem* 233:143–148

8. De Corte F, van Sluijs R, Simonitis A, Kučera J, Smodiš B, Byrne AR, De Wispelaere A, Bossus D, Frána J, Horák Z, Jačimović R (2001) Installation and calibration of Kayzero-assisted NAA in three Central European countries via a Copernicus project. *Appl Radiat Isot* 55:347–354
9. Rossbach M, Blaauw M, Bacchi M, Lin X (2007) The k₀-IAEA program. *J Radioanal Nucl Chem* 274:657–662
10. Rossbach M, Blaauw M (2006) Progress in the k₀-IAEA program. *Nucl Instr Meth A* 564:698–701
11. Robouch P, Arana G, Eguskiza M (2000) Uncertainty Budget for k₀-NAA. *J Radioanal Nucl Chem* 245:195–197
12. Blaauw M, Osorio Fernandez V, Westmeier W (1997) IAEA γ -ray spectra for testing of spectrum analysis software. *Nucl Instr Meth A* 387:410–415
13. Blaauw M, Osorio Fernandez Van Espen P, Bernasconi G, Capote Noy R, Manh Dung H, Molla NI (1997) The 1995 IAEA intercomparison of commercially available γ -ray spectrum analysis software. *Nucl Instr Meth A* 387:416–432
14. García-Toraño E, Crespo MT, Woods S, Blaauw M, Fazinic S (1999) The 1997 IAEA test spectra for alpha-particle spectrometry. *Nucl Instr Meth A* 428:330–335
15. Blaauw M, García-Toraño E, Woods S, Fazinic S (1999) The 1997 IAEA intercomparison of commercially available PC-based software for alpha-particle spectrometry. *Nucl Instr Meth A* 428:317–329
16. Blaauw M, Campbell JL, Fazinic S, Jaksic M, Orlic I, Van Espen P (2002) The 2000 IAEA intercomparison of PIXE spectrum analysis software. *Nucl Instr Meth B* 189:113–120
17. Los Arcos JM, Blaauw M, Fazinic S, Kolotov VP (2005) The 2002 IAEA test spectra for low-level γ -ray spectrometry software. *Nucl Instr Meth A* 536:189–195
18. Arnold D, Blaauw M, Fazinic S, Kolotov VP (2005) The 2002 IAEA intercomparison of software for low-level γ -ray spectrometry. *Nucl Instr Meth A* 536:196–210
19. Pessoa Barradas NP, Arstila K, Battistig G, Bianconi M, Dytlewski N, Jeynes C, Kótai E, Lulli G, Mayer M, Rauhala E, Szilágyi TM (2007) International atomic energy agency intercomparison of ion beam analysis software. *Nucl Instr Meth B* 262:281–291
20. Pessoa Barradas N, Cruz J, Fonseca M, de Jesus AP, Lagoyannis A, Manteigas V, Mayer M, Preketes-Sigalas K, Dimitriou P (2020) International atomic energy agency intercomparison of particle induced gamma-ray emission codes for bulk samples. *Nucl Instr Meth B* 468:37–47
21. Kubesova M, Kucera J (2011) Comparison of Kayzero for windows and k₀-IAEA software packages for k₀ standardization in neutron activation analysis. *Nucl Instr Meth A* 654(2011):206–212
22. Jovanović S, Dlabáč A, Mihaljević N, Vukotić P (1997) ANGLE: a PC-code for semiconductor detector efficiency calculations. *J Radioanal Nucl Chem* 218:13–20
23. Sudar S (2000) TrueCoinc, a software utility for calculation of the true coincidence correction, IAEA-TECDOC-1275:37–48
24. Gunnink R, Prindle AL (1992) Nonconventional methods for accurately calibrating germanium detectors. *J Radioanal Nucl Chem* 160:305–314
25. D'Agostino G, Di Luzio M, Oddone M (2018) An uncertainty spreadsheet for the k₀-standardisation method in Neutron Activation analysis. *J Radioanal Nucl Chem* 318:1261–1269
26. Di Luzio M, D'Agostino G, Oddone M (2020) A method to deal with correlations affecting γ counting efficiencies in analytical chemistry measurements performed by k₀-NAA. *Meas Sci Technol* 31:074006
27. Semkow TM (1990) Coincidence summing in gamma-ray spectroscopy. *Nucl Instrum Methods A* 290:437–447
28. Goorley JT (2013) Initial MCNP6 release overview, MCNP6 version 1.0 - LA-UR-13-22934
29. De Corte F (1987), The k₀-standardization Method – a move to the optimization of neutron activation analysis, habil thesis, INW, Gent, Belgium
30. HyperLab (2014) Quick start guide for main model, HyperLabs Software, Budapest, Hungary
31. (1996) KAYZERO/SOLCOI for reactor-neutron activation analysis (NAA) using the k₀-standardization method, DSM Research, Geleen, Netherlands
32. Moens L, De Donder J, Lin X, De Corte F, De Wispelaere A, Simonitis A, Hoste J (1981) Calculation of the absolute peak efficiency of gamma-ray detectors for different counting geometries. *Nucl Instr Meth* 187:451–472

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.